

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030321\
 Data File : VN066128.D
 Acq On : 3 Mar 2021 17:02
 Operator : JC/MD
 Sample : M1488-05DL 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41698DL

Quant Time: Mar 04 02:17:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030221W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 02 14:05:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	64984	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	120404	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	123155	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	48547	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.984	65	48429	48.97	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.94%
35) Dibromofluoromethane	7.547	113	37484	49.16	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.32%
50) Toluene-d8	10.055	98	148086	49.86	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.72%
62) 4-Bromofluorobenzene	12.373	95	61633	53.89	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.78%
Target Compounds						
16) Acetone	3.785	43	37394	126.51	ug/l	95
20) Methylene Chloride	4.512	84	3416	4.39	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030321\
Data File : VN066128.D
Acq On : 3 Mar 2021 17:02
Operator : JC/MD
Sample : M1488-05DL 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41698DL

Quant Time: Mar 04 02:17:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030221W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 02 14:05:08 2021
Response via : Initial Calibration

