

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030322\
 Data File : VN071149.D
 Acq On : 03 Mar 2022 13:33
 Operator : JC\MD
 Sample : PB143059TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143059TB

Quant Time: Mar 04 01:13:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	844441	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1357527	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1234943	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	411703	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	516900	53.586	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	107.180%
35) Dibromofluoromethane	8.022	113	395970	49.039	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.080%
50) Toluene-d8	10.439	98	1687223	52.032	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.060%
62) 4-Bromofluorobenzene	12.727	95	565511	48.790	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.580%
Target Compounds						
					Qvalue	
16) Acetone	4.292	43	70637	13.297	ug/l	90
20) Methylene Chloride	5.098	84	315994	32.729	ug/l	92
43) Isopropyl Acetate	8.557	43	84789	3.792	ug/l #	88
74) N-amyl acetate	12.392	43	13826	1.305	ug/l #	90
95) Naphthalene	15.503	128	23876	6.157	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030322\
Data File : VN071149.D
Acq On : 03 Mar 2022 13:33
Operator : JC\MD
Sample : PB143059TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143059TB

Quant Time: Mar 04 01:13:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

