

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030322\
 Data File : VN071152.D
 Acq On : 03 Mar 2022 14:46
 Operator : JC\MD
 Sample : PB143059ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143059ZHE#03

Quant Time: Mar 04 01:13:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	787462	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1301869	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1166156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	393667	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	487502	54.195	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.400%
35) Dibromofluoromethane	8.016	113	374724	48.392	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.780%
50) Toluene-d8	10.439	98	1602799	51.541	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.080%
62) 4-Bromofluorobenzene	12.727	95	542561	48.811	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.620%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	71713	14.476	ug/l	98
18) Methyl Acetate	4.857	43	21227	1.371	ug/l #	73
20) Methylene Chloride	5.099	84	312071	34.662	ug/l	90
43) Isopropyl Acetate	8.557	43	135416	6.314	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030322\
Data File : VN071152.D
Acq On : 03 Mar 2022 14:46
Operator : JC\MD
Sample : PB143059ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143059ZHE#03

Quant Time: Mar 04 01:13:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

