

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030322\
 Data File : VN071154.D
 Acq On : 03 Mar 2022 15:34
 Operator : JC\MD
 Sample : N1750-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2-Q1

Quant Time: Mar 04 01:14:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	827196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1357271	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1237159	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	440109	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	521669	55.208	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.420%
35) Dibromofluoromethane	8.016	113	394301	48.842	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.680%
50) Toluene-d8	10.439	98	1664373	51.337	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.680%
62) 4-Bromofluorobenzene	12.727	95	581236	50.156	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.320%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	41983	8.068	ug/l	93
18) Methyl Acetate	4.863	43	26610	1.754	ug/l	90
20) Methylene Chloride	5.098	84	18460	1.952	ug/l	88
43) Isopropyl Acetate	8.557	43	140757	6.295	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030322\
Data File : VN071154.D
Acq On : 03 Mar 2022 15:34
Operator : JC\MD
Sample : N1750-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2-Q1

Quant Time: Mar 04 01:14:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

