

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030322\
 Data File : VN071156.D
 Acq On : 03 Mar 2022 16:22
 Operator : JC\MD
 Sample : N1750-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-2-Q3

Quant Time: Mar 04 01:15:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	786800	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1290676	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1179108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	413258	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	482670	53.703	ug/l	0.00
Spiked Amount	50.000	Range 61 - 141	Recovery	= 107.400%		
35) Dibromofluoromethane	8.022	113	376101	48.991	ug/l	0.00
Spiked Amount	50.000	Range 69 - 133	Recovery	= 97.980%		
50) Toluene-d8	10.439	98	1606167	52.098	ug/l	0.00
Spiked Amount	50.000	Range 65 - 126	Recovery	= 104.200%		
62) 4-Bromofluorobenzene	12.727	95	541066	49.099	ug/l	0.00
Spiked Amount	50.000	Range 58 - 135	Recovery	= 98.200%		
Target Compounds						
					Qvalue	
16) Acetone	4.292	43	45429	9.178	ug/l	97
18) Methyl Acetate	4.857	43	26442	1.860	ug/l #	72
20) Methylene Chloride	5.098	84	18963	2.108	ug/l	95
43) Isopropyl Acetate	8.563	43	133202	6.265	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030322\
Data File : VN071156.D
Acq On : 03 Mar 2022 16:22
Operator : JC\MD
Sample : N1750-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-2-Q3

Quant Time: Mar 04 01:15:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

