



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030322\
Data File : VN071143.D
Acq On : 03 Mar 2022 09:49
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Title : SW846 8260
Last Update : Mon Feb 21 07:57:15 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1816

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.4	19413	PASS
75	95	30	60	51.1	48563	PASS
95	95	100	100	100.0	95027	PASS
96	95	5	9	7.3	6905	PASS
173	174	0.00	2	1.1	848	PASS
174	95	50	100	80.1	76112	PASS
175	174	5	9	7.2	5446	PASS
176	174	95	101	98.3	74789	PASS
177	176	5	9	6.6	4913	PASS