

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030419\
 Data File : VN054086.D
 Acq On : 5 Mar 2019 00:29
 Operator : JC/SP
 Sample : K1712-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 31 Sample Multiplier: 1


Quant Time: Mar 05 00:58:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	585580	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1091119	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1086510	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	377616	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	394339	57.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.90%	
35) Dibromofluoromethane	7.59	113	356128	50.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.94%	
50) Toluene-d8	10.09	98	1417478	53.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.94%	
62) 4-Bromofluorobenzene	12.40	95	463323	52.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.90%	
Target Compounds						
					Qvalue	
14) Allyl chloride	4.55	41	9269	0.914	ug/l #	28
16) Acetone	3.82	43	14765	Below Cal		89
18) Methyl Acetate	4.34	43	6623	1.380	ug/l #	59
20) Methylene Chloride	4.55	84	213294	31.395	ug/l	89
25) 2-Butanone	6.85	43	1811	0.672	ug/l #	47
29) Tetrahydrofuran	7.21	42	628	0.350	ug/l #	60
30) Chloroform	7.37	83	8816	0.758	ug/l	91
31) Cyclohexane	7.67	56	10785	0.577	ug/l #	1
36) 1,1-Dichloropropene	7.67	75	40860	4.027	ug/l #	48
38) Carbon Tetrachloride	7.67	117	50746	4.809	ug/l #	16
43) Isopropyl Acetate	8.18	43	35316	2.887	ug/l #	87
45) 1,2-Dichloropropane	9.27	63	1807	0.215	ug/l #	44
47) Bromodichloromethane	9.40	83	2522	0.245	ug/l #	95
48) Methyl methacrylate	9.18	41	9296	1.652	ug/l #	23
49) 1,4-Dioxane	9.25	88	30	0.394	ug/l #	25
51) 4-Methyl-2-Pentanone	10.02	43	8503	1.353	ug/l #	34
54) cis-1,3-Dichloropropene	9.90	75	62121	5.263	ug/l #	69
56) Ethyl methacrylate	10.54	69	1483	2.113	ug/l #	1
57) 1,3-Dichloropropane	10.76	76	10014	0.865	ug/l #	43
59) 2-Hexanone	10.76	43	148233	36.256	ug/l #	47
71) Bromoform	12.40	173	3683	0.641	ug/l #	1
76) 1,2,3-Trichloropropane	12.40	75	226487	37.389	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.40	75	226487	118.172	ug/l #	15
95) Naphthalene	15.13	128	2087	3.342	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030419\
Data File : VN054086.D
Acq On : 5 Mar 2019 00:29
Operator : JC/SP
Sample : K1712-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 31 Sample Multiplier: 1

Quant Time: Mar 05 00:58:19 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 28 04:05:45 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN054086.D


Acq: 5 Mar 2019 00:29

Tot Ion:168 Resp: 585580

Ion Ratio Lower Upper




168 100

99 52.7 36.9 55.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#14

Allyl chloride

Concen: 0.914 ug/l

RT: 4.55 min Scan# 865

Delta R.T. 0.23 min

Lab File: VN054086.D

Acq: 5 Mar 2019 00:29


Tgt Ion: 41 Resp: 9269

Ion Ratio Lower Upper

41 100

39 1.9 48.8 73.2#



76 0.0 30.9 46.3#




Abundance Ion 41.00 (40.70 to 41.70): VNO

Ion 39.00 (38.70 to 39.70): VNO

Ion 75.90 (75.60 to 76.60): VNO





#16

Acetone

Concen: Below Cal

RT: 3.82 min Scan# 636

Delta R.T. 0.01 min

Lab File: VN054086.D


Acq: 5 Mar 2019 00:29

Tot Ion: 43 Resp: 14765

Ion Ratio Lower Upper




43 100

58 27.9 27.6 41.4



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#18

Methyl Acetate

Concen: 1.380 ug/l

RT: 4.34 min Scan# 798

Delta R.T. 0.02 min

Lab File: VN054086.D


Acq: 5 Mar 2019 00:29

Tgt Ion: 43 Resp: 6623

Ion Ratio Lower Upper



43 100


74 4.9 20.5 30.7#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 74.00 (73.70 to 74.70): VN0






#20
Methvlene Chloride
Concen: 31.395 ug/l
RT: 4.55 min Scan# 865
Delta R.T. 0.00 min
Lab File: VN054086.D
Acq: 5 Mar 2019 00:29

Tot Ion: 84 Resp: 213294

Ion	Ratio	Lower	Upper
84	100		
49	129.3	89.7	134.5
51	40.0	27.8	41.8
86	61.9	51.3	76.9






Abundance Ion 83.90 (83.60 to 84.60): VN0

Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0


Ion 85.90 (85.60 to 86.60): VN0



#25
2-Butanone
Concen: 0.672 ug/l
RT: 6.85 min Scan# 1580
Delta R.T. 0.02 min
Lab File: VN054086.D
Acq: 5 Mar 2019 00:29



Tgt Ion: 43 Resp: 1811


Ion	Ratio	Lower	Upper
43	100		
72	0.0	22.1	33.1#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 72.00 (71.70 to 72.70): VN0





#29

Tetrahydrofuran

Concen: 0.350 ug/l

RT: 7.21 min Scan# 1692

Delta R.T. 0.00 min

Lab File: VN054086.D

Acq: 5 Mar 2019 00:29


Tot Ion: 42 Resp: 628

Ion Ratio Lower Upper

42 100

72 9.9 36.9 55.3#


71 28.0 34.6 52.0#





Abundance Ion 42.00 (41.70 to 42.70): VN054086.D (-1692) (-)

Ion 72.00 (71.70 to 72.70): VN054086.D (-1692) (-)

Ion 71.00 (70.70 to 71.70): VN054086.D (-1692) (-)



Time-->



#30

Chloroform

Concen: 0.758 ug/l

RT: 7.37 min Scan# 1741

Delta R.T. 0.00 min

Lab File: VN054086.D


Acq: 5 Mar 2019 00:29

Tgt Ion: 83 Resp: 8816

Ion Ratio Lower Upper


83 100

85 58.8 52.6 78.8





Abundance Ion 82.90 (82.60 to 83.60): VN054086.D (-1741) (-)

Ion 84.90 (84.60 to 85.60): VN054086.D (-1741) (-)




Time-->








#31
Cvclohexane
Concen: 0.577 ug/l
RT: 7.67 min Scan# 1833
Delta R.T. 0.01 min
Lab File: VN054086.D
Acq: 5 Mar 2019 00:29

Tot Ion: 56 Resp: 10785
Ion Ratio Lower Upper
56 100
69 167.7 26.0 39.0#
84 146.7 69.1 103.7#




Abundance Ion 56.00 (55.70 to 56.70): VNO
Ion 69.00 (68.70 to 69.70): VNO
Ion 84.00 (83.70 to 84.70): VNO






#33
1,2-Dichloroethane-d4
Concen: 57.448 ug/l
RT: 8.03 min Scan# 1945
Delta R.T. 0.00 min
Lab File: VN054086.D
Acq: 5 Mar 2019 00:29

Tgt Ion: 65 Resp: 394339
Ion Ratio Lower Upper
65 100
67 54.7 0.0 111.8



Abundance Ion 64.90 (64.60 to 65.60): VNO
Ion 66.90 (66.60 to 67.60): VNO





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. 0.00 min

Lab File: VN054086.D

Acq: 5 Mar 2019 00:29


Tot Ion:114 Resp: 1091119

Ion Ratio Lower Upper

114 100

63 20.8 0.0 37.4

88 15.1 0.0 30.4





Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN

8.59

Time-->



#35

Dibromofluoromethane

Concen: 50.473 ug/l

RT: 7.59 min Scan# 1809

Delta R.T. 0.00 min

Lab File: VN054086.D

Acq: 5 Mar 2019 00:29


Tgt Ion:113 Resp: 356128

Ion Ratio Lower Upper

113 100

111 101.8 81.1 121.7

192 21.2 18.8 28.2




Abundance Ion 112.80 (112.50 to 113.50): V


Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V

7.59

Time-->





#36

1,1-Dichloropropene

Concen: 4.027 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.13 min

Lab File: VN054086.D

Acq: 5 Mar 2019 00:29


Tot Ion: 75 Resp: 40860

Ion Ratio Lower Upper

75 100

110 9.1 19.0 57.0#




77 0.0 25.4 38.0#



Abundance Ion 74.90 (74.60 to 75.60): VN053925.D (-1849) (-)

Ion 109.90 (109.60 to 110.60): VN053925.D (-1849) (-)

Ion 76.90 (76.60 to 77.60): VN053925.D (-1849) (-)



#38

Carbon Tetrachloride

Concen: 4.809 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.11 min

Lab File: VN054086.D

Acq: 5 Mar 2019 00:29


Tgt Ion: 117 Resp: 50746

Ion Ratio Lower Upper

117 100

119 4.8 76.2 114.2#



121 0.0 24.6 36.8#




Abundance Ion 116.80 (116.50 to 117.50): VN053925.D (-1847) (-)

Ion 118.80 (118.50 to 119.50): VN053925.D (-1847) (-)

Ion 120.80 (120.50 to 121.50): VN053925.D (-1847) (-)





#43

Isopropyl Acetate

Concen: 2.887 ug/l

RT: 8.18 min Scan# 1991


Delta R.T. 0.01 min

Lab File: VN054086.D




Acq: 5 Mar 2019 00:29

Tot Ion: 43 Resp: 35316

Ion	Ratio	Lower	Upper
43	100		
61	16.0	17.4	26.0#
87	8.8	11.5	17.3#



Abundance Ion 43.00 (42.70 to 43.70): VN053925.D (-1969) (-)



#45

1,2-Dichloropropane

Concen: 0.215 ug/l

RT: 9.27 min Scan# 2331


Delta R.T. 0.15 min

Lab File: VN054086.D



Acq: 5 Mar 2019 00:29


Tgt Ion: 63 Resp: 1807

Ion	Ratio	Lower	Upper
63	100		
65	0.0	24.3	36.5#



Abundance Ion 62.90 (62.60 to 63.60): VN053925.D (-2261) (-)





#47

Bromodichloromethane

Concen: 0.245 ug/l

RT: 9.40 min Scan# 2373

Delta R.T. 0.00 min

Lab File: VN054086.D

Acq: 5 Mar 2019 00:29


Tot Ion: 83 Resp: 2522

Ion Ratio Lower Upper

83 100

85 63.0 51.8 77.6




127 0.0 7.4 11.0#



Abundance Ion 82.90 (82.60 to 83.60): VN054086.D (-2280) (-)

Ion 84.90 (84.60 to 85.60): VN054086.D (-2280) (-)

Ion 126.80 (126.50 to 127.50): VN054086.D (-2280) (-)



#48

Methyl methacrylate

Concen: 1.652 ug/l

RT: 9.18 min Scan# 2305

Delta R.T. -0.02 min

Lab File: VN054086.D

Acq: 5 Mar 2019 00:29


Tgt Ion: 41 Resp: 9296

Ion Ratio Lower Upper

41 100

69 0.0 74.8 112.2#



39 21.8 39.0 58.6#




Abundance Ion 41.00 (40.70 to 41.70): VN054086.D (-2217) (-)

Ion 69.00 (68.70 to 69.70): VN054086.D (-2217) (-)


Ion 39.00 (38.70 to 39.70): VN054086.D (-2217) (-)








#49
1,4-Dioxane
Concen: 0.394 ug/l
RT: 9.25 min Scan# 2326
Delta R.T. 0.05 min
Lab File: VN054086.D
Acq: 5 Mar 2019 00:29

Tot Ion: 88 Resp: 30
Ion Ratio Lower Upper
88 100
43 0.0 21.2 31.8#
58 0.0 54.2 81.2#




Abundance Ion 88.00 (87.70 to 88.70): VN0
Ion 43.00 (42.70 to 43.70): VN0
Ion 58.00 (57.70 to 58.70): VN0






#50
Toluene-d8
Concen: 53.967 ug/l
RT: 10.09 min Scan# 2587
Delta R.T. 0.00 min
Lab File: VN054086.D
Acq: 5 Mar 2019 00:29

Tgt Ion: 98 Resp: 1417478
Ion Ratio Lower Upper
98 100
100 63.4 51.9 77.9



Abundance Ion 98.00 (97.70 to 98.70): VN0
Ion 100.00 (99.70 to 100.70): VN0





#51

4-Methyl-2-Pentanone

Concen: 1.353 ug/l

RT: 10.02 min Scan# 2566

Delta R.T. 0.04 min

Lab File: VN054086.D





Acq: 5 Mar 2019 00:29

Tot Ion: 43 Resp: 8503

Ion Ratio Lower Upper

43 100

58 0.0 32.7 49.1#



#54

cis-1,3-Dichloropropene

Concen: 5.263 ug/l

RT: 9.90 min Scan# 2529

Delta R.T. 0.06 min

Lab File: VN054086.D

Acq: 5 Mar 2019 00:29




Tgt Ion: 75 Resp: 62121


Ion Ratio Lower Upper

75 100

77 0.5 25.8 38.6#

39 33.5 31.4 47.0





#56

Ethyl methacrylate

Concen: 2.113 ug/l

RT: 10.54 min Scan# 2728

Delta R.T. 0.12 min

Lab File: VN054086.D

Acq: 5 Mar 2019 00:29


Tot Ion: 69 Resp: 1483

Ion Ratio Lower Upper

69 100

41 3327.8 47.5 71.3#




39 1424.3 23.4 35.0#



Abundance Ion 69.00 (68.70 to 69.70): VN054086.D (-2728) (-)

Ion 41.00 (40.70 to 41.70): VN054086.D (-2728) (-)

Ion 39.00 (38.70 to 39.70): VN054086.D (-2728) (-)



#57

1,3-Dichloropropane

Concen: 0.865 ug/l

RT: 10.76 min Scan# 2796

Delta R.T. 0.06 min

Lab File: VN054086.D


Acq: 5 Mar 2019 00:29

Tgt Ion: 76 Resp: 10014

Ion Ratio Lower Upper



76 100


78 0.0 25.2 37.8#



Abundance Ion 75.90 (75.60 to 76.60): VN054086.D (-2796) (-)





Ion 77.90 (77.60 to 78.60): VN054086.D (-2796) (-)








#59
2-Hexanone
Concen: 36.256 ug/l
RT: 10.76 min Scan# 2796
Delta R.T. 0.01 min
Lab File: VN054086.D
Acq: 5 Mar 2019 00:29


Tot Ion: 43 Resp: 148233
Ion Ratio Lower Upper
43 100
58 19.1 29.1 87.4#



#62
4-Bromofluorobenzene
Concen: 52.945 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. -0.00 min
Lab File: VN054086.D
Acq: 5 Mar 2019 00:29


Tgt Ion: 95 Resp: 463323
Ion Ratio Lower Upper
95 100
174 87.5 0.0 188.8
176 84.4 0.0 182.6








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN054086.D
Acq: 5 Mar 2019 00:29

Tot Ion:117 Resp: 1086510
Ion Ratio Lower Upper
117 100
82 55.3 41.4 62.2
119 32.0 25.3 37.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#71
Bromoform
Concen: 0.641 ug/l
RT: 12.40 min Scan# 3306
Delta R.T. 0.28 min
Lab File: VN054086.D
Acq: 5 Mar 2019 00:29

Tgt Ion:173 Resp: 3683
Ion Ratio Lower Upper
173 100
175 656.3 24.3 72.8#
254 0.0 0.2 0.2#



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V









#72

1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.35 min Scan# 3599
Delta R.T. 0.00 min
Lab File: VN054086.D
Acq: 5 Mar 2019 00:29

Tot Ion	Ratio	Lower	Upper
152	100		
115	56.7	27.2	81.6
150	157.1	0.0	348.2




Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#76


1,2,3-Trichloropropane
Concen: 37.389 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. -0.15 min
Lab File: VN054086.D
Acq: 5 Mar 2019 00:29

Tgt Ion	Ratio	Lower	Upper
75	100		
77	1.1	0.0	0.0#



Abundance Ion 74.90 (74.60 to 75.60): V
Ion 77.00 (76.70 to 77.70): V





#81

trans-1,4-Dichloro-2-butene

Concen: 118.172 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.10 min

Lab File: VN054086.D

Acq: 5 Mar 2019 00:29


Tot Ion: 75 Resp: 226487

Ion Ratio Lower Upper

75 100

53 0.1 68.6 102.8#




89 0.0 35.3 52.9#



Abundance Ion 74.90 (74.60 to 75.60): VN054086.D (-3181) (-)

Ion 53.00 (52.70 to 53.70): VN054086.D (-3181) (-)

Ion 88.90 (88.60 to 89.60): VN054086.D (-3181) (-)



#95

Naphthalene

Concen: 3.342 ug/l

RT: 15.13 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN054086.D

Acq: 5 Mar 2019 00:29


Tgt Ion: 128 Resp: 2087

Ion Ratio Lower Upper

128 100

127 5.1 10.2 15.2#

129 3.1 8.6 12.8#



Abundance Ion 128.00 (127.70 to 128.70): VN054086.D (-4061) (-)

Ion 127.00 (126.70 to 127.70): VN054086.D (-4061) (-)

Ion 129.00 (128.70 to 129.70): VN054086.D (-4061) (-)

