

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030419\
 Data File : VN054067.D
 Acq On : 4 Mar 2019 16:00
 Operator : JC/SP
 Sample : PB117557TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117557TB

Quant Time: Mar 05 00:14:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	583996	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1068457	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1070988	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	376284	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	384533	56.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.34%	
35) Dibromofluoromethane	7.59	113	347896	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
50) Toluene-d8	10.09	98	1387217	53.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.88%	
62) 4-Bromofluorobenzene	12.40	95	455245	53.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.26%	
Target Compounds						
20) Methylene Chloride	4.55	84	40955	6.045	ug/l	90
43) Isopropyl Acetate	8.17	43	37770	3.153	ug/l #	90
95) Naphthalene	15.14	128	4921	3.521	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030419\
Data File : VN054067.D
Acq On : 4 Mar 2019 16:00
Operator : JC/SP
Sample : PB117557TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117557TB

Quant Time: Mar 05 00:14:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 28 04:05:45 2019
Response via : Initial Calibration

