

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030419\
 Data File : VN054071.D
 Acq On : 4 Mar 2019 17:47
 Operator : JC/SP
 Sample : K1688-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampled :
 JC-03-030119-B

Quant Time: Mar 05 00:23:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	563533	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1109945	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1095301	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	429664	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.03	65	394379	59.70	ug/l	0.00
Spiked Amount						
						Recovery = 119.40%
35) Dibromofluoromethane	7.59	113	358209	49.91	ug/l	0.00
Spiked Amount						
						Recovery = 99.82%
50) Toluene-d8	10.09	98	1421884	53.22	ug/l	0.00
Spiked Amount						
						Recovery = 106.44%
62) 4-Bromofluorobenzene	12.40	95	485885	54.58	ug/l	0.00
Spiked Amount						
						Recovery = 109.16%

Target Compounds

						Qvalue
20) Methylene Chloride	4.55	84	133210	20.375	ug/l	89
43) Isopropyl Acetate	8.17	43	39771	3.196	ug/l #	90
52) Toluene	10.16	92	22394	1.199	ug/l	98
68) m/p-Xylenes	11.62	106	39947	2.731	ug/l	94
69) o-Xylene	11.95	106	20269	1.410	ug/l	95
80) 1,3,5-Trimethylbenzene	12.73	105	48589	1.775	ug/l	96
84) 1,2,4-Trimethylbenzene	13.04	105	154562	5.654	ug/l	97
95) Naphthalene	15.13	128	24727	4.572	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030419\
 Data File : VN054071.D
 Acq On : 4 Mar 2019 17:47
 Operator : JC/SP
 Sample : K1688-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 JC-03-030119-B

Quant Time: Mar 05 00:23:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

