

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN030420\
 Data File : VN060342.D
 Acq On : 4 Mar 2020 12:47
 Operator : JC/MD
 Sample : PB127249TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127249TB

Quant Time: Mar 05 05:08:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	182523	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	296241	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	272675	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	123279	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	116013	49.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.08%	
35) Dibromofluoromethane	7.56	113	80905	58.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	117.68%	
50) Toluene-d8	10.08	98	365044	52.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.90%	
62) 4-Bromofluorobenzene	12.40	95	133417	50.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.14%	
Target Compounds						
					Qvalue	
16) Acetone	3.79	43	5748	7.601 ug/l		90
17) Carbon Disulfide	4.02	76	392	Below Cal	#	75
18) Methyl Acetate	4.30	43	1196	0.474 ug/l	#	52
20) Methylene Chloride	4.52	84	7974	3.839 ug/l		95
21) trans-1,2-Dichloroethene	5.00	96	100	Below Cal	#	1
25) 2-Butanone	6.82	43	1632	1.336 ug/l	#	49
29) Tetrahydrofuran	7.20	42	244	0.287 ug/l	#	33
31) Cyclohexane	7.64	56	4027	Below Cal	#	16
36) 1,1-Dichloropropene	7.64	75	14261	5.051 ug/l	#	49
37) Ethyl Acetate	6.82	43	1632	0.610 ug/l	#	68
38) Carbon Tetrachloride	7.64	117	18078	6.577 ug/l	#	16
43) Isopropyl Acetate	8.14	43	38262	8.175 ug/l	#	91
45) 1,2-Dichloropropane	9.26	63	580	0.270 ug/l	#	44
48) Methyl methacrylate	9.16	41	2427	1.137 ug/l	#	13
51) 4-Methyl-2-Pentanone	9.97	43	3707	1.397 ug/l		93
54) cis-1,3-Dichloropropene	9.89	75	16510	4.586 ug/l	#	58
57) 1,3-Dichloropropane	10.75	76	2828	0.835 ug/l	#	42
59) 2-Hexanone	10.75	43	40311	20.690 ug/l	#	53
74) N-amyl acetate	12.06	43	1013	0.254 ug/l	#	44
76) 1,2,3-Trichloropropane	12.40	75	67544	24.972 ug/l	#	100
80) 1,3,5-Trimethylbenzene	12.73	105	3940	0.512 ug/l		96
81) trans-1,4-Dichloro-2-buten	12.40	75	67544	68.935 ug/l	#	8
88) 1,4-Dichlorobenzene	13.28	146	790	Below Cal		85
93) 1,2,4-Trichlorobenzene	14.91	180	1076	0.519 ug/l		94
94) Hexachlorobutadiene	15.01	225	369	Below Cal	#	61
95) Naphthalene	15.13	128	4258	0.310 ug/l	#	93
96) 1,2,3-Trichlorobenzene	15.32	180	1220	0.339 ug/l		92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN030420\
Data File : VN060342.D
Acq On : 4 Mar 2020 12:47
Operator : JC/MD
Sample : PB127249TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127249TB

Quant Time: Mar 05 05:08:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1825

Delta R.T. -0.00 min

Lab File: VN060342.D

Acq: 4 Mar 2020 12:47

Instrument :

MSVOA_N

ClientSampleId :


PB127249TB

Tgt Ion:168 Resp: 182523

Ion Ratio Lower Upper




168 100

99 57.0 44.7 67.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0



#16

Acetone

Concen: 7.601 ug/l

RT: 3.79 min Scan# 626

Delta R.T. 0.01 min

Lab File: VN060342.D


Acq: 4 Mar 2020 12:47

Tgt Ion: 43 Resp: 5748

Ion Ratio Lower Upper



43 100


58 35.9 24.6 36.8



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0









#17
Carbon Disulfide
Concen: Below Cal
RT: 4.02 min Scan# 699
Delta R.T. -0.01 min
Lab File: VN060342.D
Acq: 4 Mar 2020 12:47




Instrument :
MSVOA_N
ClientSampleId :
PB127249TB




Tgt Ion: 76 Resp: 392
Ion Ratio Lower Upper
76 100
78 0.0 7.4 11.0#



#18
Methyl Acetate
Concen: 0.474 ug/l
RT: 4.30 min Scan# 786
Delta R.T. 0.01 min
Lab File: VN060342.D
Acq: 4 Mar 2020 12:47

Tgt Ion: 43 Resp: 1196
Ion Ratio Lower Upper
43 100
74 0.0 18.8 28.2#









#20
Methylene Chloride
Concen: 3.839 ug/l
RT: 4.52 min Scan# 854
Delta R.T. -0.00 min
Lab File: VN060342.D
Acq: 4 Mar 2020 12:47


Instrument :
MSVOA_N
ClientSampled :
PB127249TB


Tgt Ion:	84	Resp:	7974
Ion	Ratio	Lower	Upper
84	100		
49	129.6	99.3	148.9
51	43.6	30.1	45.1
86	63.2	51.1	76.7



#21
trans-1,2-Dichloroethene
Concen: Below Cal
RT: 5.00 min Scan# 1005
Delta R.T. -0.01 min
Lab File: VN060342.D
Acq: 4 Mar 2020 12:47

Tgt Ion:	96	Resp:	100
Ion	Ratio	Lower	Upper
96	100		
61	0.0	113.8	170.8#
98	0.0	48.8	73.2#





#25

2-Butanone

Concen: 1.336 ug/l

RT: 6.82 min Scan# 1568

Delta R.T. 0.02 min

Lab File: VN060342.D

Acq: 4 Mar 2020 12:47

Instrument :

MSVOA_N

ClientSampleId :


PB127249TB

Tgt Ion: 43 Resp: 1632

Ion Ratio Lower Upper

43 100

72 0.0 20.9 31.3#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 72.00 (71.70 to 72.70): VN0

6.82

500

400



300

200

100

0

Time-->



#29

Tetrahydrofuran

Concen: 0.287 ug/l

RT: 7.20 min Scan# 1687

Delta R.T. 0.02 min

Lab File: VN060342.D

Acq: 4 Mar 2020 12:47


Tgt Ion: 42 Resp: 244

Ion Ratio Lower Upper

42 100

72 0.0 34.6 52.0#

71 0.0 32.9 49.3#



Abundance Ion 42.00 (41.70 to 42.70): VN0

Ion 72.00 (71.70 to 72.70): VN0

Ion 71.00 (70.70 to 71.70): VN0

7.20

250

200


150


100

50

0

Time-->









#33
 1,2-Dichloroethane-d4
 Concen: 49.540 ug/l
 RT: 8.00 min Scan# 1937
 Delta R.T. -0.00 min
 Lab File: VN060342.D
 Acq: 4 Mar 2020 12:47




Instrument :
 MSVOA_N
 ClientSampled :
 PB127249TB


Tgt Ion: 65 Resp: 116013
 Ion Ratio Lower Upper
 65 100
 67 52.0 0.0 103.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.57 min Scan# 2113
 Delta R.T. 0.00 min
 Lab File: VN060342.D
 Acq: 4 Mar 2020 12:47

Tgt Ion: 114 Resp: 296241
 Ion Ratio Lower Upper
 114 100
 63 21.2 0.0 43.2
 88 16.1 0.0 31.6









#35
 Dibromofluoromethane
 Concen: 58.838 ug/l
 RT: 7.56 min Scan# 1801
 Delta R.T. -0.00 min
 Lab File: VN060342.D
 Acq: 4 Mar 2020 12:47




Instrument :
 MSVOA_N
 ClientSampleId :
 PB127249TB


Tgt Ion:113 Resp: 80905
 Ion Ratio Lower Upper
 113 100
 111 103.9 82.4 123.6
 192 20.5 16.2 24.2



#36
 1,1-Dichloropropene
 Concen: 5.051 ug/l
 RT: 7.64 min Scan# 1825
 Delta R.T. -0.13 min
 Lab File: VN060342.D
 Acq: 4 Mar 2020 12:47

Tgt Ion: 75 Resp: 14261
 Ion Ratio Lower Upper
 75 100
 110 7.7 17.3 52.0#
 77 0.0 24.2 36.2#





#37

Ethyl Acetate

Concen: 0.610 ug/l

RT: 6.82 min Scan# 1568

Delta R.T. -0.08 min

Lab File: VN060342.D

Acq: 4 Mar 2020 12:47

Instrument :

MSVOA_N

ClientSampleId :

PB127249TB


Tgt Ion: 43 Resp: 1632

Ion Ratio Lower Upper

43 100

61 0.0 11.1 16.7#


70 0.0 8.4 12.6#





Abundance Ion 43.00 (42.70 to 43.70): VN060342.D (-1500) (-)

Ion 60.90 (60.60 to 61.60): VN060342.D (-1500) (-)

Ion 70.00 (69.70 to 70.70): VN060342.D (-1500) (-)



Time-->



#38

Carbon Tetrachloride

Concen: 6.577 ug/l

RT: 7.64 min Scan# 1824

Delta R.T. -0.11 min

Lab File: VN060342.D

Acq: 4 Mar 2020 12:47


Tgt Ion: 117 Resp: 18078

Ion Ratio Lower Upper

117 100

119 5.7 76.7 115.1#


121 0.0 24.5 36.7#




Abundance Ion 116.80 (116.50 to 117.50): VN060342.D (-1766) (-)


Ion 118.80 (118.50 to 119.50): VN060342.D (-1766) (-)

Ion 120.80 (120.50 to 121.50): VN060342.D (-1766) (-)



Time-->





#43

Isopropyl Acetate

Concen: 8.175 ug/l

RT: 8.14 min Scan# 1980

Delta R.T. 0.00 min

Lab File: VN060342.D

Acq: 4 Mar 2020 12:47

Instrument :

MSVOA_N

ClientSampleId :

PB127249TB


Tgt Ion: 43 Resp: 38262

Ion Ratio Lower Upper

43 100

61 19.4 20.9 31.3#




87 12.5 10.6 16.0



Abundance Ion 43.00 (42.70 to 43.70): VN060342.D (-1887) (-)

Ion 61.00 (60.70 to 61.70): VN060342.D (-1887) (-)

Ion 87.00 (86.70 to 87.70): VN060342.D (-1887) (-)



#45

1,2-Dichloropropane

Concen: 0.270 ug/l

RT: 9.26 min Scan# 2327

Delta R.T. 0.15 min

Lab File: VN060342.D


Acq: 4 Mar 2020 12:47

Tgt Ion: 63 Resp: 580

Ion Ratio Lower Upper



63 100


65 0.0 24.4 36.6#



Abundance Ion 62.90 (62.60 to 63.60): VN060342.D (-2186) (-)

Ion 64.90 (64.60 to 65.60): VN060342.D (-2186) (-)









#48
Methyl methacrylate
Concen: 1.137 ug/l
RT: 9.16 min Scan# 2297
Delta R.T. -0.02 min
Lab File: VN060342.D
Acq: 4 Mar 2020 12:47




Instrument :
MSVOA_N
ClientSampled :
PB127249TB


Tgt Ion: 41 Resp: 2427
Ion Ratio Lower Upper
41 100
69 0.0 66.7 100.1#
39 0.0 50.1 75.1#



#50
Toluene-d8
Concen: 52.454 ug/l
RT: 10.08 min Scan# 2582
Delta R.T. -0.00 min
Lab File: VN060342.D
Acq: 4 Mar 2020 12:47

Tgt Ion: 98 Resp: 365044
Ion Ratio Lower Upper
98 100
100 64.7 52.2 78.2









#51
 4-Methyl-2-Pentanone
 Concen: 1.397 ug/l
 RT: 9.97 min Scan# 2548
 Delta R.T. -0.00 min
 Lab File: VN060342.D
 Acq: 4 Mar 2020 12:47




Instrument :
 MSVOA_N
 ClientSampleId :
 PB127249TB


Tgt Ion: 43 Resp: 3707
 Ion Ratio Lower Upper
 43 100
 58 34.3 30.9 46.3



#54
 cis-1,3-Dichloropropene
 Concen: 4.586 ug/l
 RT: 9.89 min Scan# 2524
 Delta R.T. 0.06 min
 Lab File: VN060342.D
 Acq: 4 Mar 2020 12:47

Tgt Ion: 75 Resp: 16510
 Ion Ratio Lower Upper
 75 100
 77 0.0 24.8 37.2#
 39 30.9 43.8 65.8#





#57

1,3-Dichloropropane

Concen: 0.835 ug/l

RT: 10.75 min Scan# 2793

Delta R.T. 0.06 min

Lab File: VN060342.D

Acq: 4 Mar 2020 12:47

Instrument :

MSVOA_N

ClientSampleId :


PB127249TB

Tgt Ion: 76 Resp: 2828

Ion Ratio Lower Upper




76 100

78 0.0 26.2 39.2#



Abundance Ion 75.90 (75.60 to 76.60): VN0

Ion 77.90 (77.60 to 78.60): VN0



#59

2-Hexanone

Concen: 20.690 ug/l

RT: 10.75 min Scan# 2793

Delta R.T. 0.01 min

Lab File: VN060342.D


Acq: 4 Mar 2020 12:47

Tgt Ion: 43 Resp: 40311

Ion Ratio Lower Upper



43 100


58 19.8 26.5 79.5#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0









#62
4-Bromofluorobenzene
Concen: 50.573 ug/l
RT: 12.40 min Scan# 3304
Delta R.T. 0.00 min
Lab File: VN060342.D
Acq: 4 Mar 2020 12:47




Instrument :
MSVOA_N
ClientSampled :
PB127249TB


Tgt Ion: 95 Resp: 133417
Ion Ratio Lower Upper
95 100
174 79.2 0.0 159.2
176 77.4 0.0 153.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.40 min Scan# 2994
Delta R.T. 0.00 min
Lab File: VN060342.D
Acq: 4 Mar 2020 12:47

Tgt Ion: 117 Resp: 272675
Ion Ratio Lower Upper
117 100
82 56.7 45.8 68.6
119 32.0 25.8 38.8





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3597

Delta R.T. 0.00 min

Lab File: VN060342.D

Acq: 4 Mar 2020 12:47

Instrument :

MSVOA_N

ClientSampleId :

PB127249TB


Tgt Ion:152 Resp: 123279

Ion Ratio Lower Upper

152 100

115 59.4 30.0 90.0




150 159.3 0.0 349.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 0.254 ug/l

RT: 12.06 min Scan# 3200

Delta R.T. 0.00 min

Lab File: VN060342.D

Acq: 4 Mar 2020 12:47

Tgt Ion: 43 Resp: 1013


Ion Ratio Lower Upper

43 100

70 0.0 32.2 48.2#

55 0.0 18.2 27.4#

61 0.0 19.2 28.8#






Abundance Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V





#76

1,2,3-Trichloropropane

Concen: 24.972 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. -0.15 min

Lab File: VN060342.D

Acq: 4 Mar 2020 12:47

Instrument :

MSVOA_N

ClientSampleId :


PB127249TB

Tgt Ion: 75 Resp: 67544

Ion Ratio Lower Upper




75 100

77 1.3 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0



#80

1,3,5-Trimethylbenzene

Concen: 0.512 ug/l

RT: 12.73 min Scan# 3408

Delta R.T. 0.00 min

Lab File: VN060342.D


Acq: 4 Mar 2020 12:47

Tgt Ion: 105 Resp: 3940

Ion Ratio Lower Upper



105 100


120 45.6 24.1 72.4



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V









#81
trans-1,4-Dichloro-2-butene
Concen: 68.935 ug/l
RT: 12.40 min Scan# 3304
Delta R.T. 0.10 min
Lab File: VN060342.D
Acq: 4 Mar 2020 12:47




Instrument :
MSVOA_N
ClientSampled :
PB127249TB


Tgt Ion: 75 Resp: 67544
Ion Ratio Lower Upper
75 100
53 0.0 73.8 110.6#
89 0.0 57.1 85.7#



#93
1,2,4-Trichlorobenzene
Concen: 0.519 ug/l
RT: 14.91 min Scan# 4086
Delta R.T. 0.01 min
Lab File: VN060342.D
Acq: 4 Mar 2020 12:47

Tgt Ion: 180 Resp: 1076
Ion Ratio Lower Upper
180 100
182 101.2 47.8 143.3
145 26.2 15.4 46.1





#94

Hexachlorobutadiene

Concen: Below Cal

RT: 15.01 min Scan# 4116

Delta R.T. 0.00 min

Lab File: VN060342.D

Acq: 4 Mar 2020 12:47

Instrument :

MSVOA_N

ClientSampled :

PB127249TB


Tgt Ion:225 Resp: 369

Ion Ratio Lower Upper

225 100

223 49.3 31.3 93.9

227 17.9 32.3 96.9#






Abundance Ion 224.70 (224.40 to 225.40): V

Ion 222.70 (222.40 to 223.40): V

Ion 226.70 (226.40 to 227.40): V

Time-->



#95

Naphthalene

Concen: 0.310 ug/l

RT: 15.13 min Scan# 4154

Delta R.T. 0.00 min

Lab File: VN060342.D

Acq: 4 Mar 2020 12:47


Tgt Ion:128 Resp: 4258

Ion Ratio Lower Upper

128 100

127 10.2 10.4 15.6#

129 8.4 8.6 13.0#






Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

Time-->








#96
 1,2,3-Trichlorobenzene
 Concen: 0.339 ug/l
 RT: 15.32 min Scan# 4212
 Delta R.T. 0.01 min
 Lab File: VN060342.D
 Acq: 4 Mar 2020 12:47

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127249TB

Tgt Ion	Ratio	Lower	Upper
180	100		
182	84.8	47.3	142.0
145	32.8	16.2	48.5



Abundance Ion 179.80 (179.50 to 180.50): V
 Ion 181.80 (181.50 to 182.50): V
 Ion 144.90 (144.60 to 145.60): V



Time-->