

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030420\
 Data File : VN060342.D
 Acq On : 4 Mar 2020 12:47
 Operator : JC/MD
 Sample : PB127249TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127249TB

Quant Time: Mar 05 06:09:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	182523	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	296241	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	272675	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	123279	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	116013	49.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.08%	
35) Dibromofluoromethane	7.56	113	80905	58.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	117.68%	
50) Toluene-d8	10.08	98	365044	52.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.90%	
62) 4-Bromofluorobenzene	12.40	95	133417	50.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.14%	
Target Compounds						
16) Acetone	3.79	43	5748	7.601	ug/l	90
20) Methylene Chloride	4.52	84	7974	3.839	ug/l	95
43) Isopropyl Acetate	8.14	43	38262	8.175	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030420\
Data File : VN060342.D
Acq On : 4 Mar 2020 12:47
Operator : JC/MD
Sample : PB127249TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127249TB

Quant Time: Mar 05 06:09:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

