


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN030420\
Data File : VN060336.D
Acq On : 4 Mar 2020 8:14
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Title : SW846 8260
Last Update : Thu Feb 27 13:52:50 2020



AutoFind: Scans 3303, 3304, 3305; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	6286	PASS
75	95	30	60	51.0	16215	PASS
95	95	100	100	100.0	31797	PASS
96	95	5	9	7.8	2484	PASS
173	174	0.00	2	1.0	249	PASS
174	95	50	100	78.8	25048	PASS
175	174	5	9	8.2	2053	PASS
176	174	95	101	100.0	25056	PASS
177	176	5	9	6.7	1677	PASS