

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030422\
 Data File : VN071161.D
 Acq On : 04 Mar 2022 11:06
 Operator : JC\MD
 Sample : VN0304MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0304MBL01

Quant Time: Mar 04 22:14:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	704776	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1129125	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	990284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	325072	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	436299	54.193	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.380%
35) Dibromofluoromethane	8.016	113	336878	50.160	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.320%
50) Toluene-d8	10.433	98	1365643	50.634	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.260%
62) 4-Bromofluorobenzene	12.727	95	399774	41.468	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	82.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030422\
Data File : VN071161.D
Acq On : 04 Mar 2022 11:06
Operator : JC\MD
Sample : VN0304MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0304MBL01

Quant Time: Mar 04 22:14:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

