

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030422\
 Data File : VN071162.D
 Acq On : 04 Mar 2022 11:29
 Operator : JC\MD
 Sample : VN0304WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0304WBL01

Quant Time: Mar 04 22:15:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	892368	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1447958	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1288936	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	441212	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	546077	53.570	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 107.140%	
35) Dibromofluoromethane	8.016	113	430558	49.992	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.980%	
50) Toluene-d8	10.433	98	1757411	50.811	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.620%	
62) 4-Bromofluorobenzene	12.727	95	589172	47.657	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 95.320%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030422\
Data File : VN071162.D
Acq On : 04 Mar 2022 11:29
Operator : JC\MD
Sample : VN0304WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0304WBL01

Quant Time: Mar 04 22:15:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

