

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030422\
 Data File : VN071165.D
 Acq On : 04 Mar 2022 12:52
 Operator : JC\MD
 Sample : N1776-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EFF-03-22

Quant Time: Mar 04 22:16:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	754553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1223482	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1075532	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	356193	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	470563	54.594	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.180%	
35) Dibromofluoromethane	8.016	113	359692	49.427	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.860%	
50) Toluene-d8	10.433	98	1481587	50.696	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 101.400%	
62) 4-Bromofluorobenzene	12.727	95	504575	48.302	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.600%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.357	59	25027	15.932	ug/l	95
16) Acetone	4.292	43	10683	2.251	ug/l	94
59) 2-Hexanone	11.080	43	12901	1.530	ug/l	95
93) 1,2,4-Trichlorobenzene	15.256	180	5387	2.788	ug/l	95
95) Naphthalene	15.503	128	15408	5.861	ug/l #	96
96) 1,2,3-Trichlorobenzene	15.698	180	6292	3.504	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030422\
Data File : VN071165.D
Acq On : 04 Mar 2022 12:52
Operator : JC\MD
Sample : N1776-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EFF-03-22

Quant Time: Mar 04 22:16:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

