

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030422\
 Data File : VN071166.D
 Acq On : 04 Mar 2022 13:16
 Operator : JC\MD
 Sample : N1776-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MID-03-22

Quant Time: Mar 04 22:16:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	868722	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1410502	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1256760	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	433227	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	535406	53.953	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	107.900%
35) Dibromofluoromethane	8.016	113	413724	49.314	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.620%
50) Toluene-d8	10.439	98	1699301	50.436	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.880%
62) 4-Bromofluorobenzene	12.727	95	565291	46.939	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.880%
Target Compounds						Qvalue
11) Tert butyl alcohol	5.351	59	35637	19.705	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030422\
Data File : VN071166.D
Acq On : 04 Mar 2022 13:16
Operator : JC\MD
Sample : N1776-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MID-03-22

Quant Time: Mar 04 22:16:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

