

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN030424\  
 Data File : VN081300.D  
 Acq On : 04 Mar 2024 13:57  
 Operator : JC\MD  
 Sample : VN0304WBSD01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :

Quant Time: Mar 05 03:20:47 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N021624W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 16 23:45:20 2024  
 Response via : Initial Calibration

Manual Integrations  
 APPROVED

Reviewed By :John  
 Carlone

| Compound                     | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min)   |
|------------------------------|--------|-------|----------|----------|--------|------------|
| -----                        |        |       |          |          |        |            |
| Internal Standards           |        |       |          |          |        | 03/05/2024 |
| 1) Pentafluorobenzene        | 8.230  | 168   | 224717   | 50.000   | ug/l   | 0.00       |
| 34) 1,4-Difluorobenzene      | 9.106  | 114   | 418718   | 50.000   | ug/l   | 0.00       |
| 63) Chlorobenzene-d5         | 11.871 | 117   | 380088   | 50.000   | ug/l   | 0.00       |
| 72) 1,4-Dichlorobenzene-d4   | 13.794 | 152   | 169769   | 50.000   | ug/l   | 0.00       |
| System Monitoring Compounds  |        |       |          |          |        | 03/05/2024 |
| 33) 1,2-Dichloroethane-d4    | 8.583  | 65    | 152063   | 55.207   | ug/l   | 0.00       |
| Spiked Amount                | 50.000 | Range | 74 - 125 | Recovery | =      | 110.420%   |
| 35) Dibromofluoromethane     | 8.171  | 113   | 133324   | 56.309   | ug/l   | 0.00       |
| Spiked Amount                | 50.000 | Range | 75 - 124 | Recovery | =      | 112.620%   |
| 50) Toluene-d8               | 10.571 | 98    | 412848   | 49.093   | ug/l   | 0.00       |
| Spiked Amount                | 50.000 | Range | 86 - 113 | Recovery | =      | 98.180%    |
| 62) 4-Bromofluorobenzene     | 12.847 | 95    | 162286   | 51.358   | ug/l   | 0.00       |
| Spiked Amount                | 50.000 | Range | 64 - 133 | Recovery | =      | 102.720%   |
| Target Compounds             |        |       |          |          |        | Qvalue     |
| 2) Dichlorodifluoromethane   | 2.124  | 85    | 58868    | 20.257   | ug/l   | 100        |
| 3) Chloromethane             | 2.365  | 50    | 56587    | 17.069   | ug/l   | 100        |
| 4) Vinyl Chloride            | 2.518  | 62    | 60377    | 18.704   | ug/l # | 83         |
| 5) Bromomethane              | 2.965  | 94    | 38369    | 16.735   | ug/l   | 97         |
| 6) Chloroethane              | 3.124  | 64    | 44637    | 20.874   | ug/l   | 99         |
| 7) Trichlorofluoromethane    | 3.501  | 101   | 107048   | 22.700   | ug/l   | 98         |
| 8) Diethyl Ether             | 3.965  | 74    | 36529    | 21.187   | ug/l   | 90         |
| 9) 1,1,2-Trichlorotrifluo... | 4.377  | 101   | 59946    | 22.802   | ug/l   | 94         |
| 10) Methyl Iodide            | 4.595  | 142   | 52566    | 17.913   | ug/l   | 98         |
| 11) Tert butyl alcohol       | 5.524  | 59    | 52513    | 103.669  | ug/l   | 98         |
| 12) 1,1-Dichloroethene       | 4.348  | 96    | 52261    | 20.565   | ug/l   | 91         |
| 13) Acrolein                 | 4.177  | 56    | 23527    | 86.656   | ug/l   | 97         |
| 14) Allyl chloride           | 5.030  | 41    | 77383    | 20.523   | ug/l # | 87         |
| 15) Acrylonitrile            | 5.730  | 53    | 161361   | 121.972  | ug/l   | 100        |
| 16) Acetone                  | 4.430  | 43    | 125568   | 115.859  | ug/l   | 95         |
| 17) Carbon Disulfide         | 4.712  | 76    | 118133   | 16.666   | ug/l # | 93         |
| 18) Methyl Acetate           | 5.024  | 43    | 74078    | 28.019   | ug/l # | 91         |
| 19) Methyl tert-butyl Ether  | 5.794  | 73    | 191549   | 22.008   | ug/l   | 98         |
| 20) Methylene Chloride       | 5.283  | 84    | 64255    | 22.404   | ug/l   | 92         |
| 21) trans-1,2-Dichloroethene | 5.794  | 96    | 57224    | 20.183   | ug/l # | 81         |
| 22) Diisopropyl ether        | 6.677  | 45    | 199633   | 23.340   | ug/l # | 92         |
| 23) Vinyl Acetate            | 6.606  | 43    | 744136   | 107.228  | ug/l # | 91         |
| 24) 1,1-Dichloroethane       | 6.571  | 63    | 120086   | 24.281   | ug/l   | 94         |
| 25) 2-Butanone               | 7.483  | 43    | 202453   | 115.628  | ug/l   | 96         |
| 26) 2,2-Dichloropropane      | 7.494  | 77    | 101858   | 21.980   | ug/l   | 96         |
| 27) cis-1,2-Dichloroethene   | 7.488  | 96    | 69635    | 21.139   | ug/l   | 90         |
| 28) Bromochloromethane       | 7.818  | 49    | 55171    | 26.908   | ug/l # | 73         |
| 29) Tetrahydrofuran          | 7.841  | 42    | 131958   | 112.899  | ug/l # | 89         |
| 30) Chloroform               | 7.965  | 83    | 126490   | 24.058   | ug/l   | 91         |
| 31) Cyclohexane              | 8.259  | 56    | 91181    | 19.172   | ug/l # | 78         |
| 32) 1,1,1-Trichloroethane    | 8.177  | 97    | 105417   | 22.015   | ug/l   | 88         |
| 36) 1,1-Dichloropropene      | 8.371  | 75    | 81614    | 20.120   | ug/l   | 95         |
| 37) Ethyl Acetate            | 7.565  | 43    | 81574    | 22.282   | ug/l   | 97         |
| 38) Carbon Tetrachloride     | 8.365  | 117   | 86820    | 21.155   | ug/l   | 94         |
| 39) Methylcyclohexane        | 9.600  | 83    | 90457    | 19.130   | ug/l   | 93         |
| 40) Benzene                  | 8.612  | 78    | 258583   | 22.023   | ug/l   | 96         |

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN030424\  
 Data File : VN081300.D  
 Acq On : 04 Mar 2024 13:57  
 Operator : JC\MD  
 Sample : VN0304WBSD01  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :

Quant Time: Mar 05 03:20:47 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N021624W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 16 23:45:20 2024  
 Response via : Initial Calibration

Manual Integrations  
 APPROVED

Reviewed By :John  
 Carlone

| Compound                      | R.T.   | QIon | Response | Conc    | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|---------|--------|----------|
| 41) Methacrylonitrile         | 7.777  | 41   | 44000    | 21.118  | ug/l   | 95       |
| 42) 1,2-Dichloroethane        | 8.671  | 62   | 95904    | 23.412  | ug/l   | 98       |
| 43) Isopropyl Acetate         | 8.688  | 43   | 144530   | 22.573  | ug/l   | 96       |
| 44) Trichloroethene           | 9.359  | 130  | 66436    | 21.244  | ug/l   | 87       |
| 45) 1,2-Dichloropropane       | 9.624  | 63   | 67217    | 22.886  | ug/l   | 100      |
| 46) Dibromomethane            | 9.712  | 93   | 47946    | 22.405  | ug/l   | 92       |
| 47) Bromodichloromethane      | 9.888  | 83   | 96114    | 23.715  | ug/l # | 96       |
| 48) Methyl methacrylate       | 9.688  | 41   | 64489    | 21.646  | ug/l   | 91       |
| 49) 1,4-Dioxane               | 9.694  | 88   | 27543    | 481.692 | ug/l # | 81       |
| 51) 4-Methyl-2-Pentanone      | 10.447 | 43   | 441293   | 119.812 | ug/l   | 97       |
| 52) Toluene                   | 10.635 | 92   | 159875   | 21.509  | ug/l   | 99       |
| 53) t-1,3-Dichloropropene     | 10.841 | 75   | 99823    | 22.164  | ug/l   | 98       |
| 54) cis-1,3-Dichloropropene   | 10.318 | 75   | 106661   | 21.919  | ug/l   | 91       |
| 55) 1,1,2-Trichloroethane     | 11.018 | 97   | 66208    | 23.159  | ug/l   | 93       |
| 56) Ethyl methacrylate        | 10.877 | 69   | 93225    | 21.336  | ug/l   | 89       |
| 57) 1,3-Dichloropropane       | 11.171 | 76   | 111191   | 22.494  | ug/l   | 99       |
| 58) 2-Chloroethyl Vinyl ether | 10.165 | 63   | 222672   | 101.570 | ug/l # | 90       |
| 59) 2-Hexanone                | 11.200 | 43   | 315565   | 108.068 | ug/l   | 96       |
| 60) Dibromochloromethane      | 11.359 | 129  | 64811    | 21.340  | ug/l   | 97       |
| 61) 1,2-Dibromoethane         | 11.471 | 107  | 65780    | 21.880  | ug/l   | 96       |
| 64) Tetrachloroethene         | 11.112 | 164  | 64240    | 24.764  | ug/l   | 90       |
| 65) Chlorobenzene             | 11.894 | 112  | 168573   | 20.492  | ug/l   | 93       |
| 66) 1,1,1,2-Tetrachloroethane | 11.959 | 131  | 62640    | 21.200  | ug/l   | 96       |
| 67) Ethyl Benzene             | 11.965 | 91   | 297119   | 20.185  | ug/l   | 98       |
| 68) m/p-Xylenes               | 12.071 | 106  | 229636   | 40.699  | ug/l   | 96       |
| 69) o-Xylene                  | 12.400 | 106  | 114069   | 20.395  | ug/l   | 93       |
| 70) Styrene                   | 12.412 | 104  | 178284   | 20.116  | ug/l   | 98       |
| 71) Bromoform                 | 12.582 | 173  | 41392    | 22.081  | ug/l # | 95       |
| 73) Isopropylbenzene          | 12.700 | 105  | 292635   | 19.701  | ug/l   | 97       |
| 74) N-amyl acetate            | 12.500 | 43   | 114863   | 18.297  | ug/l   | 93       |
| 75) 1,1,2,2-Tetrachloroethane | 12.941 | 83   | 87887    | 21.042  | ug/l   | 98       |
| 76) 1,2,3-Trichloropropane    | 12.994 | 75   | 84411m   | 21.884  | ug/l   |          |
| 77) Bromobenzene              | 12.982 | 156  | 66646    | 20.123  | ug/l   | 81       |
| 78) n-propylbenzene           | 13.041 | 91   | 348273   | 20.368  | ug/l   | 98       |
| 79) 2-Chlorotoluene           | 13.129 | 91   | 207806   | 19.813  | ug/l   | 96       |
| 80) 1,3,5-Trimethylbenzene    | 13.176 | 105  | 249782   | 20.740  | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-b... | 12.741 | 75   | 27038m   | 18.042  | ug/l   |          |
| 82) 4-Chlorotoluene           | 13.223 | 91   | 204667   | 19.957  | ug/l   | 95       |
| 83) tert-Butylbenzene         | 13.441 | 119  | 214144   | 20.564  | ug/l   | 94       |
| 84) 1,2,4-Trimethylbenzene    | 13.488 | 105  | 247634   | 20.635  | ug/l   | 95       |
| 85) sec-Butylbenzene          | 13.618 | 105  | 306748   | 21.189  | ug/l   | 96       |
| 86) p-Isopropyltoluene        | 13.729 | 119  | 243827   | 20.277  | ug/l   | 98       |
| 87) 1,3-Dichlorobenzene       | 13.735 | 146  | 122678   | 19.604  | ug/l   | 97       |
| 88) 1,4-Dichlorobenzene       | 13.817 | 146  | 125058   | 20.338  | ug/l   | 96       |
| 89) n-Butylbenzene            | 14.059 | 91   | 216615   | 20.677  | ug/l   | 97       |
| 90) Hexachloroethane          | 14.335 | 117  | 41484    | 20.871  | ug/l   | 94       |
| 91) 1,2-Dichlorobenzene       | 14.106 | 146  | 120862   | 20.025  | ug/l   | 96       |
| 92) 1,2-Dibromo-3-Chloropr... | 14.723 | 75   | 17157    | 19.757  | ug/l   | 83       |
| 93) 1,2,4-Trichlorobenzene    | 15.394 | 180  | 62380    | 18.526  | ug/l   | 97       |
| 94) Hexachlorobutadiene       | 15.506 | 225  | 25586    | 20.375  | ug/l   | 97       |
| 95) Naphthalene               | 15.647 | 128  | 219018   | 17.750  | ug/l   | 98       |
| 96) 1,2,3-Trichlorobenzene    | 15.841 | 180  | 66221    | 19.825  | ug/l   | 98       |

03/05/2024  
 Supervised By :Mahesh  
 Dadoda

03/05/2024

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN030424\  
Data File : VN081300.D  
Acq On : 04 Mar 2024 13:57  
Operator : JC\MD  
Sample : VN0304WBSD01  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Mar 05 03:20:47 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N021624W.M  
Quant Title : SW846 8260  
QLast Update : Fri Feb 16 23:45:20 2024  
Response via : Initial Calibration

Instrument :  
MSVOA\_N  
ClientSampleId :

Manual Integrations  
APPROVED


Reviewed By :John  
Carlone

| Compound                                                                         | R.T. | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------------------------------------------------------------|------|------|----------|------|-------|----------|
| -----                                                                            |      |      |          |      |       |          |
| ( # ) = qualifier out of range ( m ) = manual integration ( + ) = signals summed |      |      |          |      |       |          |

03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024





#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 8.230 min Scan# 10  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion:168 Resp: 22471  
Ion Ratio Lower Upper  
168 100  
99 60.2 59.9 89.9


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#2  
Dichlorodifluoromethane  
Concen: 20.257 ug/l  
RT: 2.124 min Scan# 29  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion: 85 Resp: 58868  
Ion Ratio Lower Upper  
85 100  
87 30.8 15.3 45.9





#3  
Chloromethane  
Concen: 17.069 ug/l  
RT: 2.365 min Scan# 70  
Delta R.T. 0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 50 Resp: 5658  
Ion Ratio Lower Upper  
50 100  
52 31.3 25.0 37.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#4  
Vinyl Chloride  
Concen: 18.704 ug/l  
RT: 2.518 min Scan# 96  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion: 62 Resp: 60377  
Ion Ratio Lower Upper  
62 100  
64 37.3 22.7 34.1#





#5  
Bromomethane  
Concen: 16.735 ug/l  
RT: 2.965 min Scan# 11  
Delta R.T. 0.011 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 94 Resp: 38369  
Ion Ratio Lower Upper  
94 100  
96 89.5 69.2 103.8


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#6  
Chloroethane  
Concen: 20.874 ug/l  
RT: 3.124 min Scan# 199  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Tgt Ion: 64 Resp: 44637  
Ion Ratio Lower Upper  
64 100  
66 32.7 26.4 39.6






#7  
Trichlorofluoromethane  
Concen: 22.700 ug/l  
RT: 3.501 min Scan# 263  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Instrument :

MSVOA\_N



ClientSampleId :



Tgt Ion:101 Resp: 107048  
Ion Ratio Lower Upper  
101 100  
103 66.2 54.4 81.6


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#8  
Diethyl Ether  
Concen: 21.187 ug/l  
RT: 3.965 min Scan# 342  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Tgt Ion: 74 Resp: 36529  
Ion Ratio Lower Upper  
74 100  
45 96.7 53.3 159.9







#9  
1,1,2-Trichlorotrifluoroethane  
Concen: 22.802 ug/l  
RT: 4.377 min Scan# 411  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Instrument :  
MSVOA\_N  
ClientSampleId :

Tgt Ion:101 Resp: 59940  
Ion Ratio Lower Upper  
101 100  
85 46.7 38.1 57.1  
151 69.3 49.1 73.7



Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#10  
Methyl Iodide  
Concen: 17.913 ug/l  
RT: 4.595 min Scan# 449  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Tgt Ion:142 Resp: 52566  
Ion Ratio Lower Upper  
142 100  
127 46.5 37.0 55.4  
141 17.8 12.5 18.7





#11

Tert butyl alcohol

Concen: 103.669 ug/l

RT: 5.524 min Scan# 605

Delta R.T. 0.012 min

Lab File: VN081300.D

Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :

Tgt Ion: 59 Resp: 5251

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 59  | 100   |       |       |
| 57  | 11.4  | 8.6   | 12.8  |



Manual Integrations  
APPROVED

Reviewed By :John  
Carlone



03/05/2024  
Supervised By :Mahesh  
Dadoda

03/05/2024



#12

1,1-Dichloroethene

Concen: 20.565 ug/l

RT: 4.348 min Scan# 407




Delta R.T. 0.006 min


Lab File: VN081300.D

Acq: 04 Mar 2024 13:57

Tgt Ion: 96 Resp: 52261


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 96  | 100   |       |       |
| 61  | 156.6 | 138.2 | 207.4 |
| 98  | 61.5  | 49.2  | 73.8  |





#13  
Acrolein  
Concen: 86.656 ug/l  
RT: 4.177 min Scan# 31  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 56 Resp: 2352  
Ion Ratio Lower Upper  
56 100  
55 67.0 55.3 82.9


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#14  
Allyl chloride  
Concen: 20.523 ug/l  
RT: 5.030 min Scan# 523  
Delta R.T. 0.012 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion: 41 Resp: 77383  
Ion Ratio Lower Upper  
41 100  
39 82.2 57.7 86.5  
76 41.5 26.8 40.2#





#15  
Acrylonitrile  
Concen: 121.972 ug/l  
RT: 5.730 min Scan# 64  
Delta R.T. 0.018 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 53 Resp: 16136  
Ion Ratio Lower Upper  
53 100  
52 81.0 64.8 97.2  
51 34.4 26.9 40.3


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#16  
Acetone  
Concen: 115.859 ug/l  
RT: 4.430 min Scan# 421  
Delta R.T. 0.012 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion: 43 Resp: 125568  
Ion Ratio Lower Upper  
43 100  
58 28.5 24.8 37.2





#17  
Carbon Disulfide  
Concen: 16.666 ug/l  
RT: 4.712 min Scan# 469  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 76 Resp: 11813  
Ion Ratio Lower Upper  
76 100  
78 10.3 6.3 9.5


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#18  
Methyl Acetate  
Concen: 28.019 ug/l  
RT: 5.024 min Scan# 522  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion: 43 Resp: 74078  
Ion Ratio Lower Upper  
43 100  
74 26.0 17.3 25.9





#19  
Methyl tert-butyl Ether  
Concen: 22.008 ug/l  
RT: 5.794 min Scan# 61  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 73 Resp: 191549  
Ion Ratio Lower Upper  
73 100  
57 23.9 18.2 27.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#20  
Methylene Chloride  
Concen: 22.404 ug/l  
RT: 5.283 min Scan# 566  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion: 84 Resp: 64255  
Ion Ratio Lower Upper  
84 100  
49 122.6 108.1 162.1  
51 36.1 32.9 49.3  
86 59.7 48.6 73.0





#21  
trans-1,2-Dichloroethene  
Concen: 20.183 ug/l  
RT: 5.794 min Scan# 61  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 96 Resp: 57224  
Ion Ratio Lower Upper  
96 100  
61 109.0 112.8 169.2  
98 60.7 51.0 76.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#22  
Diisopropyl ether  
Concen: 23.340 ug/l  
RT: 6.677 min Scan# 803  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion: 45 Resp: 199633  
Ion Ratio Lower Upper  
45 100  
43 56.4 41.8 62.8  
87 30.5 18.1 27.1  
59 12.0 9.0 13.6





#23  
 Vinyl Acetate  
 Concen: 107.228 ug/l  
 RT: 6.606 min Scan# 790  
 Delta R.T. 0.006 min  
 Lab File: VN081300.D  
 Acq: 04 Mar 2024 13:57



Instrument :  
 MSVOA\_N  
 ClientSampleId :



Tgt Ion: 43 Resp: 744130  
 Ion Ratio Lower Upper  
 43 100  
 86 12.2 7.1 10.7


Manual Integrations  
 APPROVED

Reviewed By :John  
 Carlone




03/05/2024  
 Supervised By :Mahesh  
 Dadoda



03/05/2024




#24  
 1,1-Dichloroethane  
 Concen: 24.281 ug/l  
 RT: 6.571 min Scan# 785  
 Delta R.T. -0.000 min  
 Lab File: VN081300.D  
 Acq: 04 Mar 2024 13:57




Tgt Ion: 63 Resp: 120086  
 Ion Ratio Lower Upper  
 63 100  
 98 7.9 2.9 8.6  
 100 5.4 1.9 5.7





#25  
2-Butanone  
Concen: 115.628 ug/l  
RT: 7.483 min Scan# 940  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 43 Resp: 20245  
Ion Ratio Lower Upper  
43 100  
72 26.5 19.5 29.3


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#26  
2,2-Dichloropropane  
Concen: 21.980 ug/l  
RT: 7.494 min Scan# 942  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Tgt Ion: 77 Resp: 101858  
Ion Ratio Lower Upper  
77 100  
97 21.3 11.5 34.5






#27  
cis-1,2-Dichloroethene  
Concen: 21.139 ug/l  
RT: 7.488 min Scan# 941  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Instrument :

MSVOA\_N



ClientSampleId :



Tgt Ion: 96 Resp: 6963  
Ion Ratio Lower Upper  
96 100  
61 145.7 0.0 325.4  
98 64.1 0.0 130.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#28  
Bromochloromethane  
Concen: 26.908 ug/l  
RT: 7.818 min Scan# 997  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion: 49 Resp: 55171  
Ion Ratio Lower Upper  
49 100  
129 1.4 0.0 1.6  
130 72.6 42.3 63.5#





#29  
Tetrahydrofuran  
Concen: 112.899 ug/l  
RT: 7.841 min Scan# 1000  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 42 Resp: 131958  
Ion Ratio Lower Upper  
42 100  
72 44.7 31.1 46.7  
71 43.0 28.3 42.5


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#30  
Chloroform  
Concen: 24.058 ug/l  
RT: 7.965 min Scan# 1022  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion: 83 Resp: 126490  
Ion Ratio Lower Upper  
83 100  
85 57.0 51.0 76.6





#31  
Cyclohexane  
Concen: 19.172 ug/l  
RT: 8.259 min Scan# 1072  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 56 Resp: 9118  
Ion Ratio Lower Upper  
56 100  
69 29.2 20.1 30.1  
84 93.7 56.9 85.3


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#32  
1,1,1-Trichloroethane  
Concen: 22.015 ug/l  
RT: 8.177 min Scan# 1058  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Tgt Ion: 97 Resp: 105417  
Ion Ratio Lower Upper  
97 100  
99 57.1 53.4 80.0  
61 45.5 43.8 65.8






#33  
1,2-Dichloroethane-d4  
Concen: 55.207 ug/l  
RT: 8.583 min Scan# 1127  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Instrument :

MSVOA\_N



ClientSampleId :



Tgt Ion: 65 Resp: 15206  
Ion Ratio Lower Upper  
65 100  
67 53.6 0.0 102.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 9.106 min Scan# 1216  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion:114 Resp: 418718  
Ion Ratio Lower Upper  
114 100  
63 19.5 0.0 48.0  
88 15.7 0.0 34.8





#35  
Dibromofluoromethane  
Concen: 56.309 ug/l  
RT: 8.171 min Scan# 1056  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion:113 Resp: 133324  
Ion Ratio Lower Upper  
113 100  
111 101.7 82.2 123.4  
192 16.2 12.5 18.7


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#36  
1,1-Dichloropropene  
Concen: 20.120 ug/l  
RT: 8.371 min Scan# 1091  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion: 75 Resp: 81614  
Ion Ratio Lower Upper  
75 100  
110 36.7 16.0 48.0  
77 31.5 25.8 38.6





#37  
Ethyl Acetate  
Concen: 22.282 ug/l  
RT: 7.565 min Scan# 91  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 43 Resp: 81574  
Ion Ratio Lower Upper  
43 100  
61 14.3 10.6 15.8  
70 11.4 8.0 12.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#38  
Carbon Tetrachloride  
Concen: 21.155 ug/l  
RT: 8.365 min Scan# 1090  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Tgt Ion:117 Resp: 86820  
Ion Ratio Lower Upper  
117 100  
119 98.1 73.4 110.2  
121 28.2 24.4 36.6





#39

Methylcyclohexane

Concen: 19.130 ug/l

RT: 9.600 min Scan# 1130

Delta R.T. -0.000 min


Lab File: VN081300.D

Acq: 04 Mar 2024 13:57

Instrument :

MSVOA\_N

ClientSampleId :




Tgt Ion: 83 Resp: 9045

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 83  | 100   |       |       |
| 55  | 72.7  | 65.7  | 98.5  |
| 98  | 42.1  | 33.8  | 50.6  |


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone



03/05/2024  
Supervised By :Mahesh  
Dadoda

03/05/2024



#40

Benzene

Concen: 22.023 ug/l

RT: 8.612 min Scan# 1132




Delta R.T. 0.006 min


Lab File: VN081300.D

Acq: 04 Mar 2024 13:57

Tgt Ion: 78 Resp: 258583


| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 78  | 100   |       |       |
| 77  | 23.6  | 20.4  | 30.6  |





#41  
Methacrylonitrile  
Concen: 21.118 ug/l  
RT: 7.777 min Scan# 990  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 41 Resp: 44000  
Ion Ratio Lower Upper  
41 100  
39 61.7 48.4 72.6  
67 78.1 56.8 85.2  
52 34.0 25.0 37.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#42  
1,2-Dichloroethane  
Concen: 23.412 ug/l  
RT: 8.671 min Scan# 1142  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion: 62 Resp: 95904  
Ion Ratio Lower Upper  
62 100  
98 9.6 0.0 18.0





#43  
Isopropyl Acetate  
Concen: 22.573 ug/l  
RT: 8.688 min Scan# 1145  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 43 Resp: 144530  
Ion Ratio Lower Upper  
43 100  
61 26.9 19.9 29.9  
87 13.1 9.0 13.6


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#44  
Trichloroethene  
Concen: 21.244 ug/l  
RT: 9.359 min Scan# 1259  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Tgt Ion:130 Resp: 66436  
Ion Ratio Lower Upper  
130 100  
95 92.0 0.0 210.4






#45  
1,2-Dichloropropane  
Concen: 22.886 ug/l  
RT: 9.624 min Scan# 1304  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Instrument :

MSVOA\_N


ClientSampleId :



Tgt Ion: 63 Resp: 6721  
Ion Ratio Lower Upper  
63 100  
65 32.6 26.1 39.1


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#46  
Dibromomethane  
Concen: 22.405 ug/l  
RT: 9.712 min Scan# 1319  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion: 93 Resp: 47946  
Ion Ratio Lower Upper  
93 100  
95 79.2 68.1 102.1  
174 80.5 57.2 85.8





#47  
Bromodichloromethane  
Concen: 23.715 ug/l  
RT: 9.888 min Scan# 1314  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 83 Resp: 96114  
Ion Ratio Lower Upper  
83 100  
85 68.1 52.3 78.5  
127 7.7 4.6 7.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#48  
Methyl methacrylate  
Concen: 21.646 ug/l  
RT: 9.688 min Scan# 1315  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion: 41 Resp: 64489  
Ion Ratio Lower Upper  
41 100  
69 86.9 62.2 93.2  
39 65.3 47.6 71.4





#49  
1,4-Dioxane  
Concen: 481.692 ug/l  
RT: 9.694 min Scan# 11  
Delta R.T. 0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 88 Resp: 2754  
Ion Ratio Lower Upper  
88 100  
43 23.2 30.1 45.1  
58 64.7 63.9 95.9


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#50  
Toluene-d8  
Concen: 49.093 ug/l  
RT: 10.571 min Scan# 1465  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion: 98 Resp: 412848  
Ion Ratio Lower Upper  
98 100  
100 64.5 51.4 77.0





#51  
4-Methyl-2-Pentanone  
Concen: 119.812 ug/l  
RT: 10.447 min Scan# 1444  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 43 Resp: 441293  
Ion Ratio Lower Upper  
43 100  
58 38.3 29.4 44.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#52  
Toluene  
Concen: 21.509 ug/l  
RT: 10.635 min Scan# 1476  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Tgt Ion: 92 Resp: 159875  
Ion Ratio Lower Upper  
92 100  
91 172.4 139.4 209.0







#53  
t-1,3-Dichloropropene  
Concen: 22.164 ug/l  
RT: 10.841 min Scan# 11  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Instrument :  
MSVOA\_N  
ClientSampleId :

Tgt Ion: 75 Resp: 9982  
Ion Ratio Lower Upper  
75 100  
77 30.7 25.5 38.3



Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#54  
cis-1,3-Dichloropropene  
Concen: 21.919 ug/l  
RT: 10.318 min Scan# 1422  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion: 75 Resp: 106661  
Ion Ratio Lower Upper  
75 100  
77 33.9 23.3 34.9  
39 49.1 44.6 66.8





#55  
1,1,2-Trichloroethane  
Concen: 23.159 ug/l  
RT: 11.018 min Scan# 1541  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 97 Resp: 66208  
Ion Ratio Lower Upper  
97 100  
83 87.0 72.9 109.3  
85 50.9 50.2 75.2  
99 64.0 49.6 74.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#56  
Ethyl methacrylate  
Concen: 21.336 ug/l  
RT: 10.877 min Scan# 1517  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion: 69 Resp: 93225  
Ion Ratio Lower Upper  
69 100  
41 63.0 61.0 91.6  
39 41.2 31.8 47.6





#57  
1,3-Dichloropropane  
Concen: 22.494 ug/l  
RT: 11.171 min Scan# 11119  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 76 Resp: 11119  
Ion Ratio Lower Upper  
76 100  
78 32.3 25.5 38.3


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#58  
2-Chloroethyl Vinyl ether  
Concen: 101.570 ug/l  
RT: 10.165 min Scan# 1396  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion: 63 Resp: 222672  
Ion Ratio Lower Upper  
63 100  
106 28.3 18.6 28.0#





#59  
2-Hexanone  
Concen: 108.068 ug/l  
RT: 11.200 min Scan# 1571  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 43 Resp: 31556  
Ion Ratio Lower Upper  
43 100  
58 54.7 25.9 77.8


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#60  
Dibromochloromethane  
Concen: 21.340 ug/l  
RT: 11.359 min Scan# 1599  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion:129 Resp: 64811  
Ion Ratio Lower Upper  
129 100  
127 81.7 39.4 118.2





#61  
1,2-Dibromoethane  
Concen: 21.880 ug/l  
RT: 11.471 min Scan# 107  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion:107 Resp: 65780  
Ion Ratio Lower Upper  
107 100  
109 95.1 73.4 110.2


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#62  
4-Bromofluorobenzene  
Concen: 51.358 ug/l  
RT: 12.847 min Scan# 1852  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion: 95 Resp: 162286  
Ion Ratio Lower Upper  
95 100  
174 70.6 0.0 120.4  
176 66.8 0.0 121.0





#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.871 min Scan# 1186  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion:117 Resp: 38008  
Ion Ratio Lower Upper  
117 100  
82 54.4 52.7 79.1  
119 29.5 25.3 37.9


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#64  
Tetrachloroethene  
Concen: 24.764 ug/l  
RT: 11.112 min Scan# 1557  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion:164 Resp: 64240  
Ion Ratio Lower Upper  
164 100  
166 121.1 105.8 158.6  
129 98.6 89.6 134.4  
131 97.2 84.9 127.3





#65  
Chlorobenzene  
Concen: 20.492 ug/l  
RT: 11.894 min Scan# 1100  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion:112 Resp: 16857  
Ion Ratio Lower Upper  
112 100  
114 33.4 23.7 35.5


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#66  
1,1,1,2-Tetrachloroethane  
Concen: 21.200 ug/l  
RT: 11.959 min Scan# 1701  
Delta R.T. -0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion:131 Resp: 62640  
Ion Ratio Lower Upper  
131 100  
133 98.7 46.8 140.3  
119 67.6 34.6 103.9





#67  
Ethyl Benzene  
Concen: 20.185 ug/l  
RT: 11.965 min Scan# 1702  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 91 Resp: 297119  
Ion Ratio Lower Upper  
91 100  
106 30.9 23.8 35.6


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#68  
m/p-Xylenes  
Concen: 40.699 ug/l  
RT: 12.071 min Scan# 1720  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion:106 Resp: 229636  
Ion Ratio Lower Upper  
106 100  
91 205.6 169.0 253.6





#69  
o-Xylene  
Concen: 20.395 ug/l  
RT: 12.400 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion:106 Resp: 114069  
Ion Ratio Lower Upper  
106 100  
91 212.9 112.1 336.3


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#70  
Styrene  
Concen: 20.116 ug/l  
RT: 12.412 min Scan# 1778  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion:104 Resp: 178284  
Ion Ratio Lower Upper  
104 100  
78 52.8 43.8 65.8  
103 56.4 44.2 66.2





#71  
Bromoform  
Concen: 22.081 ug/l  
RT: 12.582 min Scan# 1807  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion:173 Resp: 41392  
Ion Ratio Lower Upper  
173 100  
175 49.6 23.1 69.2  
254 0.0 0.0 0.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.794 min Scan# 2013  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion:152 Resp: 169769  
Ion Ratio Lower Upper  
152 100  
115 65.0 32.3 96.8  
150 168.5 0.0 339.8





#73  
Isopropylbenzene  
Concen: 19.701 ug/l  
RT: 12.700 min Scan# 11  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion:105 Resp: 29263  
Ion Ratio Lower Upper  
105 100  
120 26.4 12.6 37.6


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#74  
N-ethyl acetate  
Concen: 18.297 ug/l  
RT: 12.500 min Scan# 1793  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion: 43 Resp: 114863  
Ion Ratio Lower Upper  
43 100  
70 42.9 30.1 45.1  
55 25.2 17.0 25.4  
61 24.0 17.5 26.3





#75  
1,1,2,2-Tetrachloroethane  
Concen: 21.042 ug/l  
RT: 12.941 min Scan# 1868  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 83 Resp: 8788  
Ion Ratio Lower Upper  
83 100  
131 9.7 4.1 12.3  
85 65.6 32.1 96.3


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#76  
1,2,3-Trichloropropane  
Concen: 21.884 ug/l m  
RT: 12.994 min Scan# 1877  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion: 75 Resp: 84411  
Ion Ratio Lower Upper  
75 100  
77 181.2 91.0 272.8





#77  
Bromobenzene  
Concen: 20.123 ug/l  
RT: 12.982 min Scan# 1875  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion:156 Resp: 66640  
Ion Ratio Lower Upper  
156 100  
77 229.5 138.1 414.3  
158 93.7 46.5 139.5


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#78  
n-propylbenzene  
Concen: 20.368 ug/l  
RT: 13.041 min Scan# 1885  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion: 91 Resp: 348273  
Ion Ratio Lower Upper  
91 100  
120 21.7 10.3 30.9





#79  
2-Chlorotoluene  
Concen: 19.813 ug/l  
RT: 13.129 min Scan# 1900  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 91 Resp: 207800  
Ion Ratio Lower Upper  
91 100  
126 31.9 14.8 44.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#80  
1,3,5-Trimethylbenzene  
Concen: 20.740 ug/l  
RT: 13.176 min Scan# 1908  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion:105 Resp: 249782  
Ion Ratio Lower Upper  
105 100  
120 46.9 23.0 69.0





#81  
trans-1,4-Dichloro-2-butene  
Concen: 18.042 ug/l m  
RT: 12.741 min Scan# 1834  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion: 75 Resp: 27038  
Ion Ratio Lower Upper  
75 100  
53 112.4 71.4 107.2  
89 51.2 33.4 50.0


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#82  
4-Chlorotoluene  
Concen: 19.957 ug/l  
RT: 13.223 min Scan# 1916  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57


Tgt Ion: 91 Resp: 204667  
Ion Ratio Lower Upper  
91 100  
126 33.5 15.3 45.9





#83  
tert-Butylbenzene  
Concen: 20.564 ug/l  
RT: 13.441 min Scan# 1953  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion:119 Resp: 214144  
Ion Ratio Lower Upper  
119 100  
91 67.6 36.8 110.4  
134 23.7 12.4 37.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#84  
1,2,4-Trimethylbenzene  
Concen: 20.635 ug/l  
RT: 13.488 min Scan# 1961  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion:105 Resp: 247634  
Ion Ratio Lower Upper  
105 100  
120 44.9 20.9 62.7





#85  
sec-Butylbenzene  
Concen: 21.189 ug/l  
RT: 13.618 min Scan# 1983  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion:105 Resp: 306744  
Ion Ratio Lower Upper  
105 100  
134 19.1 8.6 25.8


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#86  
p-Isopropyltoluene  
Concen: 20.277 ug/l  
RT: 13.729 min Scan# 2002  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion:119 Resp: 243827  
Ion Ratio Lower Upper  
119 100  
134 25.6 12.4 37.2  
91 25.2 13.5 40.4





#87  
1,3-Dichlorobenzene  
Concen: 19.604 ug/l  
RT: 13.735 min Scan# 2003  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion:146 Resp: 122678  
Ion Ratio Lower Upper  
146 100  
111 43.1 23.9 71.7  
148 65.1 32.6 97.7


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#88  
1,4-Dichlorobenzene  
Concen: 20.338 ug/l  
RT: 13.817 min Scan# 2017  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Tgt Ion:146 Resp: 125058  
Ion Ratio Lower Upper  
146 100  
111 41.1 22.6 67.7  
148 65.3 31.9 95.5






#89  
n-Butylbenzene  
Concen: 20.677 ug/l  
RT: 14.059 min Scan# 2058  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Instrument :

MSVOA\_N



ClientSampleId :



Tgt Ion: 91 Resp: 21661  
Ion Ratio Lower Upper  
91 100  
92 51.1 26.7 80.1  
134 23.5 10.8 32.4


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#90  
Hexachloroethane  
Concen: 20.871 ug/l  
RT: 14.335 min Scan# 2105  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Tgt Ion:117 Resp: 41484  
Ion Ratio Lower Upper  
117 100  
201 55.5 30.0 90.0





#91  
1,2-Dichlorobenzene  
Concen: 20.025 ug/l  
RT: 14.106 min Scan# 2066  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Instrument :



MSVOA\_N

ClientSampleId :

Tgt Ion:146 Resp: 120862  
Ion Ratio Lower Upper  
146 100  
111 45.3 23.9 71.8  
148 65.4 31.4 94.0



Manual Integrations  
APPROVED

Reviewed By :John  
Carlone






03/05/2024  
Supervised By :Mahesh  
Dadoda


03/05/2024



#92  
1,2-Dibromo-3-Chloropropane  
Concen: 19.757 ug/l  
RT: 14.723 min Scan# 2171  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Tgt Ion: 75 Resp: 17157  
Ion Ratio Lower Upper  
75 100  
155 73.9 31.8 95.4  
157 99.7 41.1 123.4






#93  
1,2,4-Trichlorobenzene  
Concen: 18.526 ug/l  
RT: 15.394 min Scan# 21  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57

Instrument :

MSVOA\_N



ClientSampleId :



Tgt Ion:180 Resp: 62380  
Ion Ratio Lower Upper  
180 100  
182 95.6 46.6 139.7  
145 34.9 16.8 50.3


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda



03/05/2024




#94  
Hexachlorobutadiene  
Concen: 20.375 ug/l  
RT: 15.506 min Scan# 2304  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57




Tgt Ion:225 Resp: 25586  
Ion Ratio Lower Upper  
225 100  
223 66.3 33.3 99.8  
227 61.9 33.1 99.2





#95  
Naphthalene  
Concen: 17.750 ug/l  
RT: 15.647 min Scan# 219018  
Delta R.T. 0.006 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Instrument :  
MSVOA\_N  
ClientSampleId :



Tgt Ion:128 Resp: 219018  
Ion Ratio Lower Upper  
128 100  
127 13.1 10.0 15.0  
129 11.5 8.7 13.1


Manual Integrations  
APPROVED

Reviewed By :John  
Carlone




03/05/2024  
Supervised By :Mahesh  
Dadoda

03/05/2024



#96  
1,2,3-Trichlorobenzene  
Concen: 19.825 ug/l  
RT: 15.841 min Scan# 2361  
Delta R.T. -0.000 min  
Lab File: VN081300.D  
Acq: 04 Mar 2024 13:57



Tgt Ion:180 Resp: 66221  
Ion Ratio Lower Upper  
180 100  
182 95.0 46.5 139.3  
145 36.0 17.8 53.4

