

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030519\
 Data File : VN054095.D
 Acq On : 5 Mar 2019 13:27
 Operator : JC/SP
 Sample : K1726-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DSN001

Quant Time: Mar 06 02:54:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	606034	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1088910	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1031972	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	327764	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	372371	52.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.84%	
35) Dibromofluoromethane	7.59	113	350242	49.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.48%	
50) Toluene-d8	10.09	98	1364376	52.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.10%	
62) 4-Bromofluorobenzene	12.40	95	426787	48.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.74%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030519\
Data File : VN054095.D
Acq On : 5 Mar 2019 13:27
Operator : JC/SP
Sample : K1726-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DSN001

Quant Time: Mar 06 02:54:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 28 04:05:45 2019
Response via : Initial Calibration

