

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030519\
 Data File : VN054105.D
 Acq On : 5 Mar 2019 17:55
 Operator : JC/SP
 Sample : K1727-08DL 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP-1DL

Quant Time: Mar 06 03:32:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	611296	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1108810	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1083818	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	418074	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	379414	52.95	ug/l	0.00
Spiked Amount			Recovery	=	105.90%	
35) Dibromofluoromethane	7.59	113	352090	49.10	ug/l	0.00
Spiked Amount			Recovery	=	98.20%	
50) Toluene-d8	10.09	98	1414346	52.99	ug/l	0.00
Spiked Amount			Recovery	=	105.98%	
62) 4-Bromofluorobenzene	12.40	95	485739	54.62	ug/l	0.00
Spiked Amount			Recovery	=	109.24%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	2.19	62	16814	2.150	ug/l	93
27) cis-1,2-Dichloroethene	6.83	96	423182	58.261	ug/l	96
52) Toluene	10.16	92	187416	10.048	ug/l	99
67) Ethyl Benzene	11.51	91	75855	1.978	ug/l	99
68) m/p-Xylenes	11.62	106	192213	13.281	ug/l	97
69) o-Xylene	11.95	106	111779	7.860	ug/l	98
78) n-propylbenzene	12.59	91	27516	0.743	ug/l	94
80) 1,3,5-Trimethylbenzene	12.73	105	206868	7.764	ug/l	99
83) tert-Butylbenzene	12.99	119	11102	0.471	ug/l	95
84) 1,2,4-Trimethylbenzene	13.04	105	328112	12.336	ug/l	99
86) p-Isopropyltoluene	13.29	119	21656	0.796	ug/l	97
95) Naphthalene	15.13	128	32790	5.065	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030519\
Data File : VN054105.D
Acq On : 5 Mar 2019 17:55
Operator : JC/SP
Sample : K1727-08DL 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP-1DL

Quant Time: Mar 06 03:32:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 28 04:05:45 2019
Response via : Initial Calibration

