

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030520\
 Data File : VN060362.D
 Acq On : 5 Mar 2020 13:22
 Operator : JC/MD
 Sample : L1777-03
 Misc : 1.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 40681-40683

Quant Time: Mar 06 01:07:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	164216	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	262397	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	244841	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	117125	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	101237	48.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.10%	
35) Dibromofluoromethane	7.56	113	61752	50.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.40%	
50) Toluene-d8	10.08	98	323596	52.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.00%	
62) 4-Bromofluorobenzene	12.40	95	117071	50.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.20%	
Target Compounds						
3) Chloromethane	2.03	50	2886	1.101	ug/l	96
39) Methylcyclohexane	9.06	83	51153	17.349	ug/l	98
52) Toluene	10.14	92	4055502	886.993	ug/l	76
68) m/p-Xylenes	11.61	106	6709	1.975	ug/l	91
69) o-Xylene	11.94	106	3263	1.007	ug/l	92
80) 1,3,5-Trimethylbenzene	12.73	105	13286	1.818	ug/l	99
84) 1,2,4-Trimethylbenzene	13.04	105	29378	3.982	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030520\
Data File : VN060362.D
Acq On : 5 Mar 2020 13:22
Operator : JC/MD
Sample : L1777-03
Misc : 1.02g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
40681-40683

Quant Time: Mar 06 01:07:47 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

