



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030521\
 Data File : VN066130.D
 Acq On : 5 Mar 2021 13:06
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N030521W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Sat Mar 06 02:33:59 2021



AutoFind: Scans 3306, 3307, 3308; Background Corrected with Scan 3294

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	7260	PASS
75	95	30	60	50.4	21235	PASS
95	95	100	100	100.0	42112	PASS
96	95	5	9	6.3	2652	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	69.2	29152	PASS
175	174	5	9	8.4	2455	PASS
176	174	95	101	97.5	28424	PASS
177	176	5	9	6.4	1807	PASS