

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030524\
 Data File : VN081305.D
 Acq On : 05 Mar 2024 12:24
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/06/2024
 Supervised By : Mahesh Dadoda 03/06/2024

Quant Time: Mar 06 02:53:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 01:47:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	345774	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	618832	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	561947	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	252965	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	258077	51.620	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.240%			
35) Dibromofluoromethane	8.165	113	201480	53.142	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.280%			
50) Toluene-d8	10.571	98	765233	53.992	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.980%			
62) 4-Bromofluorobenzene	12.853	95	270863	54.072	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 108.140%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.124	85	246873	53.860	ug/l	93
3) Chloromethane	2.359	50	230235	46.502	ug/l	98
4) Vinyl Chloride	2.512	62	240196	46.505	ug/l	96
5) Bromomethane	2.948	94	159950	46.368	ug/l	94
6) Chloroethane	3.112	64	158348	44.636	ug/l	99
7) Trichlorofluoromethane	3.495	101	364631	47.721	ug/l	95
8) Diethyl Ether	3.953	74	131030	48.674	ug/l	84
9) 1,1,2-Trichlorotrifluo...	4.365	101	199972	45.958	ug/l	93
10) Methyl Iodide	4.589	142	204376	51.327	ug/l	97
11) Tert butyl alcohol	5.512	59	195438	223.561	ug/l	98
12) 1,1-Dichloroethene	4.336	96	190625	47.306	ug/l	92
13) Acrolein	4.177	56	248648	240.202	ug/l	100
14) Allyl chloride	5.024	41	276775	46.282	ug/l #	87
15) Acrylonitrile	5.712	53	530047	232.090	ug/l	96
16) Acetone	4.424	43	391759	221.164	ug/l	96
17) Carbon Disulfide	4.712	76	530733	46.147	ug/l	98
18) Methyl Acetate	5.018	43	249585	49.213	ug/l	92
19) Methyl tert-butyl Ether	5.800	73	652552	47.748	ug/l	95
20) Methylene Chloride	5.283	84	215652	49.261	ug/l	88
21) trans-1,2-Dichloroethene	5.789	96	207861	45.753	ug/l	98
22) Diisopropyl ether	6.671	45	660915	49.759	ug/l #	93
23) Vinyl Acetate	6.600	43	2677840	248.167	ug/l #	93
24) 1,1-Dichloroethane	6.565	63	381571	46.929	ug/l	95
25) 2-Butanone	7.483	43	684825	230.856	ug/l	94
26) 2,2-Dichloropropane	7.488	77	349529	48.640	ug/l	98
27) cis-1,2-Dichloroethene	7.488	96	234587	45.545	ug/l	91
28) Bromochloromethane	7.818	49	162638	46.748	ug/l #	72
29) Tetrahydrofuran	7.841	42	467101	230.915	ug/l	89
30) Chloroform	7.971	83	400293	47.442	ug/l	99
31) Cyclohexane	8.259	56	344053	45.454	ug/l #	80
32) 1,1,1-Trichloroethane	8.171	97	355466	47.662	ug/l	94
36) 1,1-Dichloropropene	8.377	75	298509	48.803	ug/l	97
37) Ethyl Acetate	7.559	43	294663	49.281	ug/l	100
38) Carbon Tetrachloride	8.365	117	310934	50.509	ug/l	94
39) Methylcyclohexane	9.600	83	356639	50.663	ug/l	94
40) Benzene	8.606	78	883871	48.412	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030524\
 Data File : VN081305.D
 Acq On : 05 Mar 2024 12:24
 Operator : JC\MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/06/2024
 Supervised By : Mahesh Dadoda 03/06/2024

Quant Time: Mar 06 02:53:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 01:47:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.777	41	158963	47.078	ug/l	96
42) 1,2-Dichloroethane	8.671	62	315006	48.935	ug/l	99
43) Isopropyl Acetate	8.688	43	503007	48.064	ug/l	98
44) Trichloroethene	9.353	130	226833	49.261	ug/l	98
45) 1,2-Dichloropropane	9.624	63	222150	48.436	ug/l	95
46) Dibromomethane	9.712	93	160699	49.709	ug/l	92
47) Bromodichloromethane	9.888	83	311257	49.596	ug/l #	97
48) Methyl methacrylate	9.682	41	228947	49.492	ug/l	90
49) 1,4-Dioxane	9.694	88	93743	919.404	ug/l #	70
51) 4-Methyl-2-Pentanone	10.447	43	1504059	247.512	ug/l	96
52) Toluene	10.635	92	553542	50.190	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	344842	50.784	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	369652	50.427	ug/l	96
55) 1,1,2-Trichloroethane	11.018	97	213171	48.741	ug/l	95
56) Ethyl methacrylate	10.876	69	330676	52.654	ug/l	92
57) 1,3-Dichloropropane	11.165	76	372596	48.813	ug/l	99
58) 2-Chloroethyl Vinyl ether	10.159	63	896890	265.495	ug/l #	89
59) 2-Hexanone	11.194	43	1132527	249.297	ug/l	96
60) Dibromochloromethane	11.359	129	228187	50.660	ug/l	99
61) 1,2-Dibromoethane	11.471	107	225474	49.686	ug/l	97
64) Tetrachloroethene	11.106	164	209131	47.530	ug/l	89
65) Chlorobenzene	11.894	112	572138	48.316	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.965	131	212432	49.025	ug/l	97
67) Ethyl Benzene	11.965	91	1075201	50.406	ug/l	98
68) m/p-Xylenes	12.070	106	805362	100.323	ug/l	99
69) o-Xylene	12.400	106	390584	50.203	ug/l	98
70) Styrene	12.412	104	652993	52.963	ug/l	97
71) Bromoform	12.582	173	152654	50.744	ug/l #	98
73) Isopropylbenzene	12.700	105	1044804	51.149	ug/l	99
74) N-amyl acetate	12.494	43	442279	51.603	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.941	83	294259	46.678	ug/l	99
76) 1,2,3-Trichloropropane	13.000	75	281477m	45.947	ug/l	
77) Bromobenzene	12.982	156	227939	47.064	ug/l	79
78) n-propylbenzene	13.035	91	1269446	52.542	ug/l	97
79) 2-Chlorotoluene	13.129	91	732435	49.596	ug/l	95
80) 1,3,5-Trimethylbenzene	13.176	105	876215	51.892	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.741	75	121721	56.854	ug/l	99
82) 4-Chlorotoluene	13.223	91	729122	49.985	ug/l	95
83) tert-Butylbenzene	13.441	119	735217	50.788	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	880540	51.470	ug/l	97
85) sec-Butylbenzene	13.617	105	1093407	52.210	ug/l	94
86) p-Isopropyltoluene	13.729	119	887629	52.851	ug/l	97
87) 1,3-Dichlorobenzene	13.735	146	430703	47.097	ug/l	96
88) 1,4-Dichlorobenzene	13.812	146	431296	47.185	ug/l	97
89) n-Butylbenzene	14.059	91	810020	52.592	ug/l	97
90) Hexachloroethane	14.335	117	151855	51.644	ug/l	100
91) 1,2-Dichlorobenzene	14.106	146	414373	47.226	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.717	75	66215	49.084	ug/l	84
93) 1,2,4-Trichlorobenzene	15.394	180	237878	50.720	ug/l	99
94) Hexachlorobutadiene	15.500	225	93091	48.036	ug/l	97
95) Naphthalene	15.641	128	863594	51.748	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	234345	49.272	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030524\
Data File : VN081305.D
Acq On : 05 Mar 2024 12:24
Operator : JC\MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050


Manual Integrations
APPROVED

Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024

Quant Time: Mar 06 02:53:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 01:47:06 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.224 min Scan# 1066

Delta R.T. -0.000 min

Lab File: VN081305.D

Acq: 05 Mar 2024 12:24

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:168 Resp: 345774

Ion Ratio Lower Upper

168 100





99 67.1 59.9 89.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/06/2024

Supervised By :Mahesh Dadoda 03/06/2024



#2

Dichlorodifluoromethane

Concen: 53.860 ug/l

RT: 2.124 min Scan# 29

Delta R.T. 0.000 min

Lab File: VN081305.D




Acq: 05 Mar 2024 12:24


Tgt Ion: 85 Resp: 246873

Ion Ratio Lower Upper

85 100


87 34.3 15.3 45.9





#3
Chloromethane
Concen: 46.502 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 50 Resp: 230235
Ion Ratio Lower Upper
50 100
52 32.0 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#4
Vinyl Chloride
Concen: 46.505 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion: 62 Resp: 240196
Ion Ratio Lower Upper
62 100
64 30.4 22.7 34.1





#5
Bromomethane
Concen: 46.368 ug/l
RT: 2.948 min Scan# 169
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 94 Resp: 159950
Ion Ratio Lower Upper
94 100
96 92.3 69.2 103.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#6
Chloroethane
Concen: 44.636 ug/l
RT: 3.112 min Scan# 197
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion: 64 Resp: 158348
Ion Ratio Lower Upper
64 100
66 32.6 26.4 39.6





#7
 Trichlorofluoromethane
 Concen: 47.721 ug/l
 RT: 3.495 min Scan# 262
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050



Tgt Ion:101 Resp: 36463
 Ion Ratio Lower Upper
 101 100
 103 63.7 54.4 81.6




Manual Integrations
 APPROVED


Reviewed By :John Carlone 03/06/2024
 Supervised By :Mahesh Dadoda 03/06/2024



#8
 Diethyl Ether
 Concen: 48.674 ug/l
 RT: 3.953 min Scan# 340
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24

Tgt Ion: 74 Resp: 131030
 Ion Ratio Lower Upper
 74 100
 45 89.6 53.3 159.9









#9
1,1,2-Trichlorotrifluoroethane
Concen: 45.958 ug/l
RT: 4.365 min Scan# 410
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:101 Resp: 199972
Ion Ratio Lower Upper
101 100
85 47.0 38.1 57.1
151 70.9 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#10
Methyl Iodide
Concen: 51.327 ug/l
RT: 4.589 min Scan# 448
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Tgt Ion:142 Resp: 204376
Ion Ratio Lower Upper
142 100
127 44.4 37.0 55.4
141 14.7 12.5 18.7









#11
Tert butyl alcohol
Concen: 223.561 ug/l
RT: 5.512 min Scan# 605
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 59 Resp: 195438
Ion Ratio Lower Upper
59 100
57 10.1 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#12
1,1-Dichloroethene
Concen: 47.306 ug/l
RT: 4.336 min Scan# 405
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion: 96 Resp: 190625
Ion Ratio Lower Upper
96 100
61 157.9 138.2 207.4
98 60.3 49.2 73.8





#13
Acrolein
Concen: 240.202 ug/l
RT: 4.177 min Scan# 31
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 56 Resp: 248648
Ion Ratio Lower Upper
56 100
55 68.7 55.3 82.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#14
Allyl chloride
Concen: 46.282 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion: 41 Resp: 276775
Ion Ratio Lower Upper
41 100
39 82.8 57.7 86.5
76 41.5 26.8 40.2#





#15
Acrylonitrile
Concen: 232.090 ug/l
RT: 5.712 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 53 Resp: 53004
Ion Ratio Lower Upper
53 100
52 83.5 64.8 97.2
51 36.6 26.9 40.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#16
Acetone
Concen: 221.164 ug/l
RT: 4.424 min Scan# 420
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Tgt Ion: 43 Resp: 391759
Ion Ratio Lower Upper
43 100
58 28.8 24.8 37.2





#17

Carbon Disulfide

Concen: 46.147 ug/l

RT: 4.712 min Scan# 469

Delta R.T. 0.000 min

Lab File: VN081305.D

Acq: 05 Mar 2024 12:24

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 53073

Ion Ratio Lower Upper

76 100





78 8.8 6.3 9.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/06/2024

Supervised By :Mahesh Dadoda 03/06/2024



#18

Methyl Acetate

Concen: 49.213 ug/l

RT: 5.018 min Scan# 521

Delta R.T. 0.000 min

Lab File: VN081305.D




Acq: 05 Mar 2024 12:24


Tgt Ion: 43 Resp: 249585

Ion Ratio Lower Upper

43 100

74 25.6 17.3 25.9









#19
Methyl tert-butyl Ether
Concen: 47.748 ug/l
RT: 5.800 min Scan# 654
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 73 Resp: 652552
Ion Ratio Lower Upper
73 100
57 20.4 18.2 27.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#20
Methylene Chloride
Concen: 49.261 ug/l
RT: 5.283 min Scan# 566
Delta R.T. -0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion: 84 Resp: 215652
Ion Ratio Lower Upper
84 100
49 117.9 108.1 162.1
51 35.2 32.9 49.3
86 55.8 48.6 73.0





#21
trans-1,2-Dichloroethene
Concen: 45.753 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 96 Resp: 20786
Ion Ratio Lower Upper
96 100
61 137.7 112.8 169.2
98 64.5 51.0 76.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#22
Diisopropyl ether
Concen: 49.759 ug/l
RT: 6.671 min Scan# 802
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Tgt Ion: 45 Resp: 660915
Ion Ratio Lower Upper
45 100
43 55.4 41.8 62.8
87 30.6 18.1 27.1#
59 10.9 9.0 13.6









#23
 Vinyl Acetate
 Concen: 248.167 ug/l
 RT: 6.600 min Scan# 790
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Tgt Ion: 43 Resp: 2677840
 Ion Ratio Lower Upper
 43 100
 86 11.4 7.1 10.7




Manual Integrations
 APPROVED


Reviewed By :John Carlone 03/06/2024
 Supervised By :Mahesh Dadoda 03/06/2024



#24
 1,1-Dichloroethane
 Concen: 46.929 ug/l
 RT: 6.565 min Scan# 784
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24


Tgt Ion: 63 Resp: 381571
 Ion Ratio Lower Upper
 63 100
 98 7.6 2.9 8.6
 100 5.3 1.9 5.7





#25
2-Butanone
Concen: 230.856 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050




Tgt Ion: 43 Resp: 684825
Ion Ratio Lower Upper
43 100
72 27.6 19.5 29.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024




#26
2,2-Dichloropropane
Concen: 48.640 ug/l
RT: 7.488 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Tgt Ion: 77 Resp: 349529
Ion Ratio Lower Upper
77 100
97 22.0 11.5 34.5





#27
 cis-1,2-Dichloroethene
 Concen: 45.545 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050



Tgt Ion: 96 Resp: 23458
 Ion Ratio Lower Upper
 96 100
 61 146.8 0.0 325.4
 98 66.1 0.0 130.4




Manual Integrations
 APPROVED


Reviewed By :John Carlone 03/06/2024
 Supervised By :Mahesh Dadoda 03/06/2024



#28
 Bromochloromethane
 Concen: 46.748 ug/l
 RT: 7.818 min Scan# 997
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24


Tgt Ion: 49 Resp: 162638
 Ion Ratio Lower Upper
 49 100
 129 2.3 0.0 1.6#
 130 72.8 42.3 63.5#





#29
Tetrahydrofuran
Concen: 230.915 ug/l
RT: 7.841 min Scan# 1001
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 42 Resp: 46710
Ion Ratio Lower Upper
42 100
72 45.4 31.1 46.7
71 41.9 28.3 42.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#30
Chloroform
Concen: 47.442 ug/l
RT: 7.971 min Scan# 1023
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion: 83 Resp: 400293
Ion Ratio Lower Upper
83 100
85 64.3 51.0 76.6





#31
Cyclohexane
Concen: 45.454 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 56 Resp: 34405
Ion Ratio Lower Upper
56 100
69 27.8 20.1 30.1
84 92.1 56.9 85.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#32
1,1,1-Trichloroethane
Concen: 47.662 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Tgt Ion: 97 Resp: 355466
Ion Ratio Lower Upper
97 100
99 63.4 53.4 80.0
61 49.3 43.8 65.8





#33
 1,2-Dichloroethane-d4
 Concen: 51.620 ug/l
 RT: 8.577 min Scan# 1126
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion: 65 Resp: 25807
 Ion Ratio Lower Upper
 65 100
 67 51.3 0.0 102.4

Manual Integrations

APPROVED




Reviewed By :John Carlone 03/06/2024


Supervised By :Mahesh Dadoda 03/06/2024



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 9.106 min Scan# 1216
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24


Tgt Ion:114 Resp: 618832
 Ion Ratio Lower Upper
 114 100
 63 21.2 0.0 48.0
 88 15.7 0.0 34.8





#35
Dibromofluoromethane
Concen: 53.142 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:113 Resp: 201480
Ion Ratio Lower Upper
113 100
111 103.4 82.2 123.4
192 16.8 12.5 18.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#36
1,1-Dichloropropene
Concen: 48.803 ug/l
RT: 8.377 min Scan# 1092
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Tgt Ion: 75 Resp: 298509
Ion Ratio Lower Upper
75 100
110 35.2 16.0 48.0
77 31.6 25.8 38.6









#37
Ethyl Acetate
Concen: 49.281 ug/l
RT: 7.559 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 29466
Ion Ratio Lower Upper
43 100
61 13.5 10.6 15.8
70 10.0 8.0 12.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#38
Carbon Tetrachloride
Concen: 50.509 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Tgt Ion:117 Resp: 310934
Ion Ratio Lower Upper
117 100
119 98.3 73.4 110.2
121 29.7 24.4 36.6





#39

Methylcyclohexane

Concen: 50.663 ug/l

RT: 9.600 min Scan# 11300

Delta R.T. 0.000 min

Lab File: VN081305.D

Acq: 05 Mar 2024 12:24

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:	83	Resp:	356639
Ion Ratio	Lower	Upper	
83	100		
55	79.0	65.7	98.5
98	49.3	33.8	50.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/06/2024

Supervised By :Mahesh Dadoda 03/06/2024



#40

Benzene

Concen: 48.412 ug/l




RT: 8.606 min Scan# 1131


Delta R.T. 0.000 min

Lab File: VN081305.D

Acq: 05 Mar 2024 12:24

Tgt Ion:	78	Resp:	883871
Ion Ratio	Lower	Upper	
78	100		
77	23.9	20.4	30.6









#41
Methacrylonitrile
Concen: 47.078 ug/l
RT: 7.777 min Scan# 990
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 41 Resp: 15896
Ion Ratio Lower Upper
41 100
39 59.8 48.4 72.6
67 76.0 56.8 85.2
52 33.4 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#42
1,2-Dichloroethane
Concen: 48.935 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Tgt Ion: 62 Resp: 315006
Ion Ratio Lower Upper
62 100
98 9.5 0.0 18.0









#43
Isopropyl Acetate
Concen: 48.064 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 50300
Ion Ratio Lower Upper
43 100
61 25.6 19.9 29.9
87 13.1 9.0 13.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#44
Trichloroethene
Concen: 49.261 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Tgt Ion:130 Resp: 226833
Ion Ratio Lower Upper
130 100
95 103.1 0.0 210.4





#45

1,2-Dichloropropane

Concen: 48.436 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.000 min

Lab File: VN081305.D

Acq: 05 Mar 2024 12:24

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 63 Resp: 222150

Ion Ratio Lower Upper

63 100





65 29.7 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/06/2024

Supervised By :Mahesh Dadoda 03/06/2024



#46

Dibromomethane

Concen: 49.709 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.000 min

Lab File: VN081305.D

Acq: 05 Mar 2024 12:24




Tgt Ion: 93 Resp: 160699


Ion Ratio Lower Upper

93 100

95 82.0 68.1 102.1

174 82.0 57.2 85.8





#47

Bromodichloromethane

Concen: 49.596 ug/l

RT: 9.888 min Scan# 1349

Delta R.T. 0.000 min

Lab File: VN081305.D

Acq: 05 Mar 2024 12:24

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 31125





Ion	Ratio	Lower	Upper
83	100		
85	63.5	52.3	78.5
127	9.1	4.6	7.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/06/2024

Supervised By :Mahesh Dadoda 03/06/2024



#48

Methyl methacrylate

Concen: 49.492 ug/l

RT: 9.682 min Scan# 1314




Delta R.T. 0.000 min


Lab File: VN081305.D

Acq: 05 Mar 2024 12:24

Tgt Ion: 41 Resp: 228947

Ion	Ratio	Lower	Upper
41	100		
69	87.7	62.2	93.2
39	66.3	47.6	71.4









#49
1,4-Dioxane
Concen: 919.404 ug/l
RT: 9.694 min Scan# 1316
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 88 Resp: 93743
Ion Ratio Lower Upper
88 100
43 0.0 30.1 45.1
58 66.4 63.9 95.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#50
Toluene-d8
Concen: 53.992 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Tgt Ion: 98 Resp: 765233
Ion Ratio Lower Upper
98 100
100 65.9 51.4 77.0









#51
4-Methyl-2-Pentanone
Concen: 247.512 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 1504059
Ion Ratio Lower Upper
43 100
58 39.0 29.4 44.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#52
Toluene
Concen: 50.190 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion: 92 Resp: 553542
Ion Ratio Lower Upper
92 100
91 172.0 139.4 209.0





#53
t-1,3-Dichloropropene
Concen: 50.784 ug/l
RT: 10.835 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050




Tgt Ion: 75 Resp: 344842
Ion Ratio Lower Upper
75 100
77 31.5 25.5 38.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024




#54
cis-1,3-Dichloropropene
Concen: 50.427 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Tgt Ion: 75 Resp: 369652
Ion Ratio Lower Upper
75 100
77 33.0 23.3 34.9
39 53.8 44.6 66.8





#55
1,1,2-Trichloroethane
Concen: 48.741 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 97 Resp: 21317
Ion Ratio Lower Upper
97 100
83 89.3 72.9 109.3
85 57.3 50.2 75.2
99 67.8 49.6 74.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#56
Ethyl methacrylate
Concen: 52.654 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion: 69 Resp: 330676
Ion Ratio Lower Upper
69 100
41 65.8 61.0 91.6
39 39.1 31.8 47.6





#57
1,3-Dichloropropane
Concen: 48.813 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 76 Resp: 372590
Ion Ratio Lower Upper
76 100
78 32.7 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#58
2-Chloroethyl Vinyl ether
Concen: 265.495 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion: 63 Resp: 896890
Ion Ratio Lower Upper
63 100
106 28.5 18.6 28.0#





#59
2-Hexanone
Concen: 249.297 ug/l
RT: 11.194 min Scan# 1571
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 43 Resp: 113252
Ion Ratio Lower Upper
43 100
58 54.4 25.9 77.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#60
Dibromochloromethane
Concen: 50.660 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion:129 Resp: 228187
Ion Ratio Lower Upper
129 100
127 78.3 39.4 118.2





#61
1,2-Dibromoethane
Concen: 49.686 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050




Tgt Ion:107 Resp: 225474
Ion Ratio Lower Upper
107 100
109 94.6 73.4 110.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024




#62
4-Bromofluorobenzene
Concen: 54.072 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24



Tgt Ion: 95 Resp: 270863
Ion Ratio Lower Upper
95 100
174 72.7 0.0 120.4
176 70.0 0.0 121.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.871 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN081305.D

Acq: 05 Mar 2024 12:24

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:117 Resp: 56194





Ion	Ratio	Lower	Upper
117	100		
82	55.1	52.7	79.1
119	32.4	25.3	37.9

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/06/2024

Supervised By :Mahesh Dadoda 03/06/2024



#64

Tetrachloroethene

Concen: 47.530 ug/l

RT: 11.106 min Scan# 1556




Delta R.T. 0.000 min


Lab File: VN081305.D

Acq: 05 Mar 2024 12:24

Tgt Ion:164 Resp: 209131

Ion	Ratio	Lower	Upper
164	100		
166	120.4	105.8	158.6
129	96.3	89.6	134.4
131	97.7	84.9	127.3









#65
Chlorobenzene
Concen: 48.316 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:112 Resp: 572138
Ion Ratio Lower Upper
112 100
114 31.4 23.7 35.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#66
1,1,1,2-Tetrachloroethane
Concen: 49.025 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion:131 Resp: 212432
Ion Ratio Lower Upper
131 100
133 95.8 46.8 140.3
119 66.0 34.6 103.9





#67
Ethyl Benzene
Concen: 50.406 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050




Tgt Ion: 91 Resp: 107520
Ion Ratio Lower Upper
91 100
106 28.7 23.8 35.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024




#68
m/p-Xylenes
Concen: 100.323 ug/l
RT: 12.070 min Scan# 1720
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Tgt Ion:106 Resp: 805362
Ion Ratio Lower Upper
106 100
91 209.2 169.0 253.6





#69
o-Xylene
Concen: 50.203 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050




Tgt Ion:106 Resp: 390584
Ion Ratio Lower Upper
106 100
91 221.6 112.1 336.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024




#70
Styrene
Concen: 52.963 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Tgt Ion:104 Resp: 652993
Ion Ratio Lower Upper
104 100
78 51.3 43.8 65.8
103 54.7 44.2 66.2





#71
Bromoform
Concen: 50.744 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050




Tgt Ion:173 Resp: 152654
Ion Ratio Lower Upper
173 100
175 47.2 23.1 69.2
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Tgt Ion:152 Resp: 252965
Ion Ratio Lower Upper
152 100
115 65.5 32.3 96.8
150 174.2 0.0 339.8





#73
Isopropylbenzene
Concen: 51.149 ug/l
RT: 12.700 min Scan# 1827
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion:105 Resp: 1044804
Ion Ratio Lower Upper
105 100
120 25.5 12.6 37.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#74
N-ethyl acetate
Concen: 51.603 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion: 43 Resp: 442279
Ion Ratio Lower Upper
43 100
70 41.5 30.1 45.1
55 24.0 17.0 25.4
61 23.8 17.5 26.3





#75
 1,1,2,2-Tetrachloroethane
 Concen: 46.678 ug/l
 RT: 12.941 min Scan# 1868
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050



Tgt Ion: 83 Resp: 294259
 Ion Ratio Lower Upper
 83 100
 131 10.2 4.1 12.3
 85 64.4 32.1 96.3




Manual Integrations
 APPROVED


Reviewed By :John Carlone 03/06/2024
 Supervised By :Mahesh Dadoda 03/06/2024



#76
 1,2,3-Trichloropropane
 Concen: 45.947 ug/l m
 RT: 13.000 min Scan# 1878
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24

Tgt Ion: 75 Resp: 281477
 Ion Ratio Lower Upper
 75 100
 77 183.0 91.0 272.8





#77

Bromobenzene

Concen: 47.064 ug/l

RT: 12.982 min Scan# 1875

Delta R.T. 0.000 min

Lab File: VN081305.D

Acq: 05 Mar 2024 12:24

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050

Tgt Ion:156 Resp: 227939

Ion Ratio Lower Upper

156 100

77 225.9 138.1 414.3





158 96.9 46.5 139.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/06/2024

Supervised By :Mahesh Dadoda 03/06/2024



#78

n-propylbenzene

Concen: 52.542 ug/l

RT: 13.035 min Scan# 1884

Delta R.T. 0.000 min

Lab File: VN081305.D




Acq: 05 Mar 2024 12:24


Tgt Ion: 91 Resp: 1269446

Ion Ratio Lower Upper

91 100

120 21.8 10.3 30.9









#79
2-Chlorotoluene
Concen: 49.596 ug/l
RT: 13.129 min Scan# 1900
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 732435
Ion Ratio Lower Upper
91 100
126 32.2 14.8 44.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#80
1,3,5-Trimethylbenzene
Concen: 51.892 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion:105 Resp: 876215
Ion Ratio Lower Upper
105 100
120 46.5 23.0 69.0





#81
trans-1,4-Dichloro-2-butene
Concen: 56.854 ug/l
RT: 12.741 min Scan# 1834
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050



Tgt Ion: 75 Resp: 12172
Ion Ratio Lower Upper
75 100
53 89.4 71.4 107.2
89 39.2 33.4 50.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#82
4-Chlorotoluene
Concen: 49.985 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion: 91 Resp: 729122
Ion Ratio Lower Upper
91 100
126 33.1 15.3 45.9





#83
tert-Butylbenzene
Concen: 50.788 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24




Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050




Tgt Ion:119 Resp: 73521
Ion Ratio Lower Upper
119 100
91 70.7 36.8 110.4
134 26.1 12.4 37.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024




#84
1,2,4-Trimethylbenzene
Concen: 51.470 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24



Tgt Ion:105 Resp: 880540
Ion Ratio Lower Upper
105 100
120 43.4 20.9 62.7









#85
 sec-Butylbenzene
 Concen: 52.210 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050

Tgt Ion:105 Resp: 109340
 Ion Ratio Lower Upper
 105 100
 134 19.6 8.6 25.8




Manual Integrations
 APPROVED


Reviewed By :John Carlone 03/06/2024
 Supervised By :Mahesh Dadoda 03/06/2024



#86
 p-Isopropyltoluene
 Concen: 52.851 ug/l
 RT: 13.729 min Scan# 2002
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24


Tgt Ion:119 Resp: 887629
 Ion Ratio Lower Upper
 119 100
 134 25.7 12.4 37.2
 91 25.0 13.5 40.4





#87
 1,3-Dichlorobenzene
 Concen: 47.097 ug/l
 RT: 13.735 min Scan# 2003
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICCC050




Tgt Ion:146 Resp: 43070
 Ion Ratio Lower Upper
 146 100
 111 43.2 23.9 71.7
 148 63.6 32.6 97.7

Manual Integrations
 APPROVED



Reviewed By :John Carlone 03/06/2024
 Supervised By :Mahesh Dadoda 03/06/2024




#88
 1,4-Dichlorobenzene
 Concen: 47.185 ug/l
 RT: 13.812 min Scan# 2016
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24



Tgt Ion:146 Resp: 431296
 Ion Ratio Lower Upper
 146 100
 111 41.8 22.6 67.7
 148 62.2 31.9 95.5









#89
n-Butylbenzene
Concen: 52.592 ug/l
RT: 14.059 min Scan# 2058
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 810020
Ion Ratio Lower Upper
91 100
92 52.1 26.7 80.1
134 24.1 10.8 32.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#90
Hexachloroethane
Concen: 51.644 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Tgt Ion:117 Resp: 151855
Ion Ratio Lower Upper
117 100
201 60.1 30.0 90.0









#91
1,2-Dichlorobenzene
Concen: 47.226 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24

Instrument :
MSVOA_N
ClientSampleId :
VSTDICCC050

Tgt Ion:146 Resp: 41437
Ion Ratio Lower Upper
146 100
111 43.5 23.9 71.8
148 62.6 31.4 94.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 03/06/2024
Supervised By :Mahesh Dadoda 03/06/2024



#92
1,2-Dibromo-3-Chloropropane
Concen: 49.084 ug/l
RT: 14.717 min Scan# 2170
Delta R.T. 0.000 min
Lab File: VN081305.D
Acq: 05 Mar 2024 12:24


Tgt Ion: 75 Resp: 66215
Ion Ratio Lower Upper
75 100
155 77.2 31.8 95.4
157 95.7 41.1 123.4





#93
 1,2,4-Trichlorobenzene
 Concen: 50.720 ug/l
 RT: 15.394 min Scan# 2285
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24

Instrument : MSVOA_N
 ClientSampleId : VSTDICCC050






Tgt Ion:180 Resp: 237878


Ion	Ratio	Lower	Upper
180	100		
182	94.1	46.6	139.7
145	33.7	16.8	50.3

Manual Integrations
 APPROVED

Reviewed By :John Carlone 03/06/2024
 Supervised By :Mahesh Dadoda 03/06/2024






#94
 Hexachlorobutadiene
 Concen: 48.036 ug/l
 RT: 15.500 min Scan# 2303
 Delta R.T. 0.000 min
 Lab File: VN081305.D
 Acq: 05 Mar 2024 12:24



Tgt Ion:225 Resp: 93091

Ion	Ratio	Lower	Upper
225	100		
223	64.0	33.3	99.8
227	63.3	33.1	99.2





#95

Naphthalene

Concen: 51.748 ug/l

RT: 15.641 min Scan# 2327

Delta R.T. 0.000 min

Lab File: VN081305.D

Acq: 05 Mar 2024 12:24

Instrument :

MSVOA_N

ClientSampleId :

VSTDICCC050





Tgt Ion:	128	Resp:	863594
Ion Ratio	Lower	Upper	
128	100		
127	12.9	10.0	15.0
129	10.9	8.7	13.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 03/06/2024

Supervised By :Mahesh Dadoda 03/06/2024



#96

1,2,3-Trichlorobenzene

Concen: 49.272 ug/l

RT: 15.841 min Scan# 2361

Delta R.T. 0.000 min

Lab File: VN081305.D

Acq: 05 Mar 2024 12:24

Tgt Ion:	180	Resp:	234345
Ion Ratio	Lower	Upper	
180	100		
182	96.2	46.5	139.3
145	35.8	17.8	53.4

