

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030619\
 Data File : VN054124.D
 Acq On : 6 Mar 2019 18:39
 Operator : JC/SP
 Sample : K1755-29
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-5

Quant Time: Mar 07 07:11:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 06:25:49 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	586673	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1038086	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1004086	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	368904	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.03	65	361948	51.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.74%	
35) Dibromofluoromethane	7.59	113	335980	49.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.66%	
50) Toluene-d8	10.09	98	1322613	51.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.32%	
62) 4-Bromofluorobenzene	12.40	95	448434	53.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.60%	

Target Compounds


						Qvalue
16) Acetone	3.82	43	41873	19.296	ug/l	100
25) 2-Butanone	6.84	43	7862	2.723	ug/l	95
39) Methylcyclohexane	9.08	83	2017	0.179	ug/l #	78
40) Benzene	8.04	78	120796	4.190	ug/l	98
41) Methacrylonitrile	7.18	41	59	0.016	ug/l #	100
42) 1,2-Dichloroethane	8.04	62	1154	0.126	ug/l #	74
44) Trichloroethene	8.83	130	371	0.048	ug/l	13
46) Dibromomethane	9.21	93	29	0.006	ug/l #	2
48) Methyl methacrylate	9.23	41	240	0.045	ug/l #	41
52) Toluene	10.16	92	127425	7.461	ug/l	100
53) t-1,3-Dichloropropene	10.39	75	252	0.027	ug/l #	44
54) cis-1,3-Dichloropropene	9.84	75	233	0.021	ug/l #	52
55) 1,1,2-Trichloroethane	10.43	97	30	0.005	ug/l #	15
56) Ethyl methacrylate	10.44	69	210	0.028	ug/l #	1
57) 1,3-Dichloropropane	10.70	76	30	0.003	ug/l #	42
60) Dibromochloromethane	10.63	129	1127	0.164	ug/l #	11
64) Tetrachloroethene	10.63	164	1327	0.161	ug/l	96
65) Chlorobenzene	11.43	112	1487	0.071	ug/l	92
66) 1,1,1,2-Tetrachloroethane	11.50	131	62	0.008	ug/l #	8
67) Ethyl Benzene	11.51	91	50203	1.436	ug/l	99
68) m/p-Xylenes	11.62	106	108666	8.235	ug/l	99
69) o-Xylene	11.95	106	52205	4.111	ug/l	98
70) Styrene	11.96	104	19785	0.982	ug/l #	68
80) 1,3,5-Trimethylbenzene	12.73	105	25665	1.076	ug/l	100
84) 1,2,4-Trimethylbenzene	13.04	105	72580	3.065	ug/l	99
95) Naphthalene	15.13	128	39446	7.194	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030619\
Data File : VN054124.D
Acq On : 6 Mar 2019 18:39
Operator : JC/SP
Sample : K1755-29
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-5

Quant Time: Mar 07 07:11:05 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 06:25:49 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39

Instrument :

MSVOA_N

ClientSampled :


COMP-5

Tot Ion:168 Resp: 586673

Ion Ratio Lower Upper

168 100

99 52.6 41.1 61.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.67

200000



150000

100000

50000

0

Time-->



#16

Acetone

Concen: 19.296 ug/l

RT: 3.82 min Scan# 635

Delta R.T. 0.00 min

Lab File: VN054124.D


Acq: 6 Mar 2019 18:39

Tgt Ion: 43 Resp: 41873

Ion Ratio Lower Upper

43 100

58 33.1 26.5 39.7



Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 58.00 (57.70 to 58.70): VNO

3.82


15000


10000

5000

0

Time-->









#25
2-Butanone
Concen: 2.723 ug/l
RT: 6.84 min Scan# 1577
Delta R.T. 0.01 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39




Instrument :
MSVOA_N
ClientSampled :
COMP-5


Tot Ion: 43 Resp: 7862
Ion Ratio Lower Upper
43 100
72 22.7 20.2 30.2



#33
1,2-Dichloroethane-d4
Concen: 51.371 ug/l
RT: 8.03 min Scan# 1945
Delta R.T. 0.00 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39

Tgt Ion: 65 Resp: 361948
Ion Ratio Lower Upper
65 100
67 55.0 0.0 109.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. 0.00 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39

Instrument :

MSVOA_N

ClientSampleId :

COMP-5


Tot Ion:114 Resp: 1038086

Ion Ratio Lower Upper

114 100

63 20.7 0.0 42.6




88 15.7 0.0 31.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 49.833 ug/l

RT: 7.59 min Scan# 1809

Delta R.T. 0.00 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39


Tgt Ion:113 Resp: 335980

Ion Ratio Lower Upper

113 100

111 102.7 82.1 123.1



192 21.6 17.4 26.0




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V







#39
Methvlcvclohexane
Concen: 0.179 ug/l
RT: 9.08 min Scan# 2273
Delta R.T. 0.00 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39

Instrument :
MSVOA_N
ClientSampled :
COMP-5



Tot Ion: 83 Resp: 2017
Ion Ratio Lower Upper
83 100
55 54.4 62.1 93.1#
98 36.6 36.0 54.0



Abundance Ion 83.00 (82.70 to 83.70): VN0
Ion 55.00 (54.70 to 55.70): VN0
Ion 98.00 (97.70 to 98.70): VN0




Time-->




#40
Benzene
Concen: 4.190 ug/l
RT: 8.04 min Scan# 1949
Delta R.T. 0.00 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39


Tgt Ion: 78 Resp: 120796
Ion Ratio Lower Upper
78 100
77 22.6 18.7 28.1






Abundance Ion 78.00 (77.70 to 78.70): VN0
Ion 77.00 (76.70 to 77.70): VN0



Time-->









#41
Methacrylonitrile
Concen: 0.016 ug/l
RT: 7.18 min Scan# 1683
Delta R.T. -0.00 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39


Instrument :
MSVOA_N
ClientSampled :
COMP-5


Tot Ion:	41	Resp:	59
Ion	Ratio	Lower	Upper
41	100		
39	161.0	0.0	0.0#
67	0.0	0.0	0.0
52	0.0	0.0	0.0



#42
1,2-Dichloroethane
Concen: 0.126 ug/l
RT: 8.04 min Scan# 1950
Delta R.T. -0.08 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39

Tgt Ion:	62	Resp:	1154
Ion	Ratio	Lower	Upper
62	100		
98	0.0	0.0	18.6





#44

Trichloroethene

Concen: 0.048 ug/l

RT: 8.83 min Scan# 2196

Delta R.T. 0.00 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39

Instrument :

MSVOA_N

ClientSampled :


COMP-5

Tot Ion:130 Resp: 371

Ion Ratio Lower Upper


130 100

95 10.7 0.0 189.4





Abundance Ion 129.80 (129.50 to 130.50): VN054124.D (-2179) (-)

Ion 94.90 (94.60 to 95.60): VN054124.D (-2179) (-)



Time-->



#46

Dibromomethane

Concen: 0.006 ug/l

RT: 9.21 min Scan# 2313

Delta R.T. 0.00 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39


Tgt Ion: 93 Resp: 29

Ion Ratio Lower Upper

93 100

95 0.0 66.1 99.1#


174 0.0 86.1 129.1#




Abundance Ion 92.80 (92.50 to 93.50): VN054124.D (-2219) (-)


Ion 94.80 (94.50 to 95.50): VN054124.D (-2219) (-)

Ion 173.70 (173.40 to 174.40): VN054124.D (-2219) (-)



Time-->









#48
Methvl methacrylate
Concen: 0.045 ug/l
RT: 9.23 min Scan# 2318
Delta R.T. 0.03 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39




Instrument :
MSVOA_N
ClientSampled :
COMP-5


Tot Ion: 41 Resp: 240
Ion Ratio Lower Upper
41 100
69 0.0 71.0 106.4#
39 54.6 43.7 65.5



#50
Toluene-d8
Concen: 51.659 ug/l
RT: 10.09 min Scan# 2587
Delta R.T. 0.00 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39

Tgt Ion: 98 Resp: 1322613
Ion Ratio Lower Upper
98 100
100 63.5 51.7 77.5





#52

Toluene

Concen: 7.461 ug/l

RT: 10.16 min Scan# 2607

Delta R.T. 0.00 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39

Instrument :

MSVOA_N

ClientSampleId :





COMP-5

Tot Ion: 92 Resp: 127425

Ion Ratio Lower Upper

92 100

91 173.5 139.0 208.6



#53

t-1,3-Dichloropropene

Concen: 0.027 ug/l

RT: 10.39 min Scan# 2681

Delta R.T. 0.01 min

Lab File: VN054124.D




Acq: 6 Mar 2019 18:39


Tgt Ion: 75 Resp: 252

Ion Ratio Lower Upper

75 100

77 0.0 24.9 37.3#





#54

cis-1,3-Dichloropropene

Concen: 0.021 ug/l

RT: 9.84 min Scan# 2509

Delta R.T. 0.00 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39

Instrument :

MSVOA_N

ClientSampled :

COMP-5





Tot Ion: 75 Resp: 233

Ion Ratio Lower Upper

75 100

77 0.0 25.4 38.0#

39 18.4 36.2 54.2#



#55

1,1,2-Trichloroethane

Concen: 0.005 ug/l

RT: 10.43 min Scan# 2693

Delta R.T. -0.13 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39

Tgt Ion: 97 Resp: 30




Ion Ratio Lower Upper


97 100

83 0.0 69.8 104.6#

85 0.0 45.0 67.4#

99 0.0 49.8 74.6#





#56

Ethyl methacrylate

Concen: 0.028 ug/l

RT: 10.44 min Scan# 2694

Delta R.T. 0.01 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39

Instrument :

MSVOA_N

ClientSampled :

COMP-5





Tot Ion: 69 Resp: 210

Ion Ratio Lower Upper

69 100

41 214.8 53.4 80.2#

39 0.0 25.6 38.4#



#57

1,3-Dichloropropane

Concen: 0.003 ug/l

RT: 10.70 min Scan# 2777

Delta R.T. -0.00 min

Lab File: VN054124.D




Acq: 6 Mar 2019 18:39


Tgt Ion: 76 Resp: 30

Ion Ratio Lower Upper

76 100

78 0.0 25.9 38.9#





#60

Dibromochloromethane

Concen: 0.164 ug/l

RT: 10.63 min Scan# 2756

Delta R.T. -0.27 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39

Instrument :

MSVOA_N

ClientSampleId :


COMP-5

Tot Ion:129 Resp: 1127

Ion Ratio Lower Upper

129 100

127 0.0 38.8 116.4#



Abundance Ion 128.80 (128.50 to 129.50): V

Ion 126.80 (126.50 to 127.50): V

10.63



600

400

200

0

Time-->



#62

4-Bromofluorobenzene

Concen: 53.798 ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.00 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39


Tgt Ion: 95 Resp: 448434

Ion Ratio Lower Upper

95 100

174 88.3 0.0 173.2

176 84.6 0.0 168.8



Abundance Ion 94.90 (94.60 to 95.60): V

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V


300000


200000

100000

0

Time-->









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39

Instrument :
MSVOA_N
ClientSampled :
COMP-5

Tot Ion:117 Resp: 1004086
Ion Ratio Lower Upper
117 100
82 55.5 43.4 65.2
119 32.0 26.0 39.0




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#64
Tetrachloroethene
Concen: 0.161 ug/l
RT: 10.63 min Scan# 2756
Delta R.T. 0.00 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39

Tgt Ion:164 Resp: 1327
Ion Ratio Lower Upper
164 100
166 121.0 101.3 151.9
129 88.1 73.2 109.8
131 91.7 70.8 106.2



Abundance Ion 163.80 (163.50 to 164.50): V
Ion 165.80 (165.50 to 166.50): V
Ion 128.80 (128.50 to 129.50): V
Ion 130.80 (130.50 to 131.50): V









#65
Chlorobenzene
Concen: 0.071 ug/l
RT: 11.43 min Scan# 3002
Delta R.T. -0.01 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39




Instrument :
MSVOA_N
ClientSampled :
COMP-5


Tot Ion:112 Resp: 1487
Ion Ratio Lower Upper
112 100
114 35.9 25.4 38.0



#66
1,1,1,2-Tetrachloroethane
Concen: 0.008 ug/l
RT: 11.50 min Scan# 3025
Delta R.T. -0.01 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39

Tgt Ion:131 Resp: 62
Ion Ratio Lower Upper
131 100
133 0.0 47.6 142.9#
119 0.0 33.8 101.3#









#67
Ethyl Benzene
Concen: 1.436 ug/l
RT: 11.51 min Scan# 3028
Delta R.T. 0.00 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39




Instrument :
MSVOA_N
ClientSampled :
COMP-5


Tot Ion: 91 Resp: 50203
Ion Ratio Lower Upper
91 100
106 31.4 24.6 36.8



#68
m/p-Xylenes
Concen: 8.235 ug/l
RT: 11.62 min Scan# 3061
Delta R.T. -0.01 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39

Tgt Ion: 106 Resp: 108666
Ion Ratio Lower Upper
106 100
91 202.0 163.2 244.8









#69
o-Xylene
Concen: 4.111 ug/l
RT: 11.95 min Scan# 3164
Delta R.T. 0.00 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39

Instrument :
MSVOA_N
ClientSampled :
COMP-5

Tot Ion:106 Resp: 52205
Ion Ratio Lower Upper
106 100
91 216.9 107.1 321.3




Abundance Ion 106.00 (105.70 to 106.70): V
Ion 91.00 (90.70 to 91.70): VN






#70
Styrene
Concen: 0.982 ug/l
RT: 11.96 min Scan# 3169
Delta R.T. 0.00 min
Lab File: VN054124.D
Acq: 6 Mar 2019 18:39

Tgt Ion:104 Resp: 19785
Ion Ratio Lower Upper
104 100
78 74.9 39.4 59.0#
103 75.6 44.7 67.1#



Abundance Ion 104.00 (103.70 to 104.70): V
Ion 78.00 (77.70 to 78.70): VN
Ion 103.00 (102.70 to 103.70): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.34 min Scan# 3598

Delta R.T. 0.00 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39

Instrument :

MSVOA_N

ClientSampleId :

COMP-5


Tot Ion:152 Resp: 368904

Ion Ratio Lower Upper

152 100

115 57.6 28.5 85.6




150 158.7 0.0 343.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#80

1,3,5-Trimethylbenzene

Concen: 1.076 ug/l

RT: 12.73 min Scan# 3408

Delta R.T. 0.00 min

Lab File: VN054124.D


Acq: 6 Mar 2019 18:39

Tgt Ion:105 Resp: 25665

Ion Ratio Lower Upper



105 100


120 48.9 24.6 73.8



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#84

1,2,4-Trimethylbenzene

Concen: 3.065 ug/l

RT: 13.04 min Scan# 3504

Delta R.T. 0.00 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39

Instrument :

MSVOA_N

ClientSampled :





COMP-5

Tot Ion:105 Resp: 72580

Ion Ratio Lower Upper

105 100

120 44.3 22.6 67.7



#95

Naphthalene

Concen: 7.194 ug/l

RT: 15.13 min Scan# 4155

Delta R.T. 0.00 min

Lab File: VN054124.D

Acq: 6 Mar 2019 18:39

Tgt Ion:128 Resp: 39446

Ion Ratio Lower Upper

128 100

127 13.1 10.4 15.6

129 11.4 8.6 12.8

