

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030620\
 Data File : VN060382.D
 Acq On : 6 Mar 2020 15:51
 Operator : JC/MD
 Sample : PB127296ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127296ZHE#01

Quant Time: Mar 07 03:56:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	185624	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	304516	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	288070	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	136969	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	119857	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
35) Dibromofluoromethane	7.57	113	87160	61.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.32%	
50) Toluene-d8	10.08	98	376191	52.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.18%	
62) 4-Bromofluorobenzene	12.40	95	143202	52.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.62%	
Target Compounds						
20) Methylene Chloride	4.52	84	3505	1.659	ug/l	91
43) Isopropyl Acetate	8.14	43	30176	6.272	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030620\
Data File : VN060382.D
Acq On : 6 Mar 2020 15:51
Operator : JC/MD
Sample : PB127296ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127296ZHE#01

Quant Time: Mar 07 03:56:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

