

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030620\
 Data File : VN060384.D
 Acq On : 6 Mar 2020 16:36
 Operator : JC/MD
 Sample : PB127296ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127296ZHE#03

Quant Time: Mar 07 04:12:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	196781	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	323114	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	302400	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	141075	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	125672	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
35) Dibromofluoromethane	7.57	113	89732	59.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.66%	
50) Toluene-d8	10.08	98	397665	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
62) 4-Bromofluorobenzene	12.40	95	148299	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
Target Compounds						
20) Methylene Chloride	4.53	84	3635	1.623	ug/l	97
43) Isopropyl Acetate	8.14	43	29341	5.748	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030620\
Data File : VN060384.D
Acq On : 6 Mar 2020 16:36
Operator : JC/MD
Sample : PB127296ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127296ZHE#03

Quant Time: Mar 07 04:12:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

