

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN030620\
 Data File : VN060376.D
 Acq On : 6 Mar 2020 13:35
 Operator : JC/MD
 Sample : L1828-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 07 03:27:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	193278	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	316886	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	287760	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	125853	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	122671	49.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.94%	
35) Dibromofluoromethane	7.57	113	88790	60.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	120.74%	
50) Toluene-d8	10.08	98	387799	52.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.18%	
62) 4-Bromofluorobenzene	12.40	95	138158	48.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.92%	


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN030620\
Data File : VN060376.D
Acq On : 6 Mar 2020 13:35
Operator : JC/MD
Sample : L1828-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Mar 07 03:27:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.64 min Scan# 1826

Delta R.T. 0.00 min

Lab File: VN060376.D

Acq: 6 Mar 2020 13:35

Instrument :

MSVOA_N

ClientSampleId :


STORAGE-BLANK-WATER-REF-5

Tot Ion:168 Resp: 193278

Ion Ratio Lower Upper




168 100

99 54.6 44.7 67.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0



#33

1,2-Dichloroethane-d4

Concen: 49.468 ug/l

RT: 8.00 min Scan# 1938

Delta R.T. 0.00 min

Lab File: VN060376.D


Acq: 6 Mar 2020 13:35

Tgt Ion: 65 Resp: 122671

Ion Ratio Lower Upper



65 100


67 53.2 0.0 103.8



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.57 min Scan# 2113

Delta R.T. 0.00 min

Lab File: VN060376.D

Acq: 6 Mar 2020 13:35

Instrument :

MSVOA_N

ClientSampleId :

STORAGE-BLANK-WATER-REF-5


Tot Ion:114 Resp: 316886

Ion Ratio Lower Upper

114 100

63 21.5 0.0 43.2




88 16.3 0.0 31.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): V

Ion 87.90 (87.60 to 88.60): V



#35

Dibromofluoromethane

Concen: 60.365 ug/l

RT: 7.57 min Scan# 1802

Delta R.T. 0.00 min

Lab File: VN060376.D

Acq: 6 Mar 2020 13:35


Tgt Ion:113 Resp: 88790

Ion Ratio Lower Upper

113 100

111 102.7 82.4 123.6



192 19.8 16.2 24.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 52.093 ug/l

RT: 10.08 min Scan# 2583

Delta R.T. 0.00 min

Lab File: VN060376.D

Acq: 6 Mar 2020 13:35

Instrument :

MSVOA_N

ClientSampleId :





STORAGE-BLANK-WATER-REF-5

Tot Ion: 98 Resp: 387799

Ion Ratio Lower Upper

98 100

100 64.6 52.2 78.2



#62

4-Bromofluorobenzene

Concen: 48.958 ug/l

RT: 12.40 min Scan# 3304

Delta R.T. 0.00 min

Lab File: VN060376.D

Acq: 6 Mar 2020 13:35




Tgt Ion: 95 Resp: 138158


Ion Ratio Lower Upper

95 100

174 79.5 0.0 159.2

176 77.4 0.0 153.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.40 min Scan# 2994
Delta R.T. 0.00 min
Lab File: VN060376.D
Acq: 6 Mar 2020 13:35

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Tot Ion:117 Resp: 287760
Ion Ratio Lower Upper
117 100
82 57.0 45.8 68.6
119 32.4 25.8 38.8




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.34 min Scan# 3597
Delta R.T. 0.00 min
Lab File: VN060376.D
Acq: 6 Mar 2020 13:35

Tgt Ion:152 Resp: 125853
Ion Ratio Lower Upper
152 100
115 60.6 30.0 90.0
150 159.2 0.0 349.6



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

