

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030624\
 Data File : VN081314.D
 Acq On : 06 Mar 2024 10:03
 Operator : JC\MD
 Sample : VN0306MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0306MBL01

Quant Time: Mar 07 00:21:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	308653	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	566498	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	489798	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	192537	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	223226	50.019	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.040%
35) Dibromofluoromethane	8.171	113	177113	51.030	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.060%
50) Toluene-d8	10.571	98	659473	50.829	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.660%
62) 4-Bromofluorobenzene	12.853	95	205866	44.893	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	89.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030624\
Data File : VN081314.D
Acq On : 06 Mar 2024 10:03
Operator : JC\MD
Sample : VN0306MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0306MBL01

Quant Time: Mar 07 00:21:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

