

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030624\
 Data File : VN081324.D
 Acq On : 06 Mar 2024 14:14
 Operator : JC\MD
 Sample : PB159453ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB159453ZHE#05

Quant Time: Mar 07 00:26:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	289143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	546549	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	486318	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	180966	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	219430	52.487	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.980%	
35) Dibromofluoromethane	8.171	113	171182	51.122	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.240%	
50) Toluene-d8	10.571	98	657089	52.493	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.980%	
62) 4-Bromofluorobenzene	12.853	95	204513	46.226	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 92.460%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	111182	75.060	ug/l	97
20) Methylene Chloride	5.277	84	52174	13.672	ug/l	91
37) Ethyl Acetate	7.559	43	6816	1.291	ug/l #	91
43) Isopropyl Acetate	8.694	43	311244	33.674	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030624\
Data File : VN081324.D
Acq On : 06 Mar 2024 14:14
Operator : JC\MD
Sample : PB159453ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB159453ZHE#05

Quant Time: Mar 07 00:26:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

