

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030624\
 Data File : VN081327.D
 Acq On : 06 Mar 2024 15:26
 Operator : JC\MD
 Sample : P1672-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/07/2024
 Supervised By : Mahesh Dadoda 03/07/2024

Quant Time: Mar 07 00:27:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	311432	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	570241	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	516570	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	193955	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	232730	51.684	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 103.360%			
35) Dibromofluoromethane	8.171	113	180127	51.558	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.120%			
50) Toluene-d8	10.571	98	695485	53.252	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.500%			
62) 4-Bromofluorobenzene	12.853	95	217714	47.165	ug/l	0.00
Spiked Amount 50.000	Range 64 - 133		Recovery = 94.340%			
Target Compounds						
						Qvalue
16) Acetone	4.430	43	38403	24.071	ug/l	95
20) Methylene Chloride	5.271	84	14234m	2.852	ug/l	
37) Ethyl Acetate	7.577	43	6751	1.225	ug/l	95
43) Isopropyl Acetate	8.688	43	320801	33.266	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030624\
Data File : VN081327.D
Acq On : 06 Mar 2024 15:26
Operator : JC\MD
Sample : P1672-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5

Quant Time: Mar 07 00:27:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 03/07/2024
Supervised By :Mahesh Dadoda 03/07/2024

