

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030624\
 Data File : VN081328.D
 Acq On : 06 Mar 2024 15:51
 Operator : JC\MD
 Sample : P1672-16
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-13

Quant Time: Mar 07 00:28:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	294164	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	555295	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	499861	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	185376	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	218202	51.302	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.600%		
35) Dibromofluoromethane	8.165	113	174762	51.369	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.740%		
50) Toluene-d8	10.570	98	671521	52.801	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 105.600%		
62) 4-Bromofluorobenzene	12.853	95	206925	46.035	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 92.060%		
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	41015	27.217	ug/l	95
20) Methylene Chloride	5.271	84	13748	2.935	ug/l	90
37) Ethyl Acetate	7.577	43	6889	1.284	ug/l #	90
43) Isopropyl Acetate	8.688	43	344511	36.686	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030624\
Data File : VN081328.D
Acq On : 06 Mar 2024 15:51
Operator : JC\MD
Sample : P1672-16
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-13

Quant Time: Mar 07 00:28:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

