

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030624\
 Data File : VN081329.D
 Acq On : 06 Mar 2024 16:15
 Operator : JC\MD
 Sample : P1673-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RT-2321

Quant Time: Mar 07 00:28:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	292913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	548017	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	495631	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	187300	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	223727	52.826	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.660%	
35) Dibromofluoromethane	8.171	113	174873	52.084	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.160%	
50) Toluene-d8	10.571	98	657648	52.397	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.800%	
62) 4-Bromofluorobenzene	12.853	95	206910	46.643	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 93.280%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	43715	29.133	ug/l	99
20) Methylene Chloride	5.289	84	14609	3.187	ug/l #	72
43) Isopropyl Acetate	8.688	43	323625	34.919	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030624\
Data File : VN081329.D
Acq On : 06 Mar 2024 16:15
Operator : JC\MD
Sample : P1673-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RT-2321

Quant Time: Mar 07 00:28:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

