

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030624\
 Data File : VN081332.D
 Acq On : 06 Mar 2024 17:27
 Operator : JC\MD
 Sample : P1672-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Mar 07 00:29:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	294204	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	552518	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	489726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	177703	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	217423	51.112	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.220%		
35) Dibromofluoromethane	8.171	113	174064	51.421	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.840%		
50) Toluene-d8	10.571	98	679246	53.677	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 107.360%		
62) 4-Bromofluorobenzene	12.853	95	206679	46.211	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 92.420%		
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	33383	22.150	ug/l	91
20) Methylene Chloride	5.277	84	13340	2.823	ug/l	92
37) Ethyl Acetate	7.571	43	6616	1.239	ug/l #	88
43) Isopropyl Acetate	8.688	43	366814	39.257	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030624\
Data File : VN081332.D
Acq On : 06 Mar 2024 17:27
Operator : JC\MD
Sample : P1672-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Mar 07 00:29:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

