

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030625\
 Data File : VN085908.D
 Acq On : 06 Mar 2025 19:07
 Operator : JC\MD
 Sample : PB166989ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166989ZHE#10

Quant Time: Mar 07 00:44:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	134004	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	253221	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	242067	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	85535	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	104429	60.109	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 120.220%			
35) Dibromofluoromethane	8.165	113	88976	53.698	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.400%			
50) Toluene-d8	10.565	98	324638	53.850	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.700%			
62) 4-Bromofluorobenzene	12.847	95	95735	48.199	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 96.400%			
Target Compounds						
16) Acetone	4.424	43	29499	53.662	ug/l	93
43) Isopropyl Acetate	8.688	43	64833m	18.802	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030625\
Data File : VN085908.D
Acq On : 06 Mar 2025 19:07
Operator : JC\MD
Sample : PB166989ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166989ZHE#10

Quant Time: Mar 07 00:44:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

