

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030625\
 Data File : VN085911.D
 Acq On : 06 Mar 2025 20:20
 Operator : JC\MD
 Sample : PB166989ZHE#13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166989ZHE#13

Quant Time: Mar 07 00:45:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	132357	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	248858	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	245749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	90842	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	104592	60.952	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 121.900%		
35) Dibromofluoromethane	8.165	113	89789	55.139	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 110.280%		
50) Toluene-d8	10.565	98	318697	53.792	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 107.580%		
62) 4-Bromofluorobenzene	12.847	95	98078	50.244	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 100.480%		
Target Compounds						
						Qvalue
16) Acetone	4.436	43	31833	58.628	ug/l	95
20) Methylene Chloride	5.265	84	3417	1.957	ug/l #	81
43) Isopropyl Acetate	8.683	43	64673m	19.093	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030625\
Data File : VN085911.D
Acq On : 06 Mar 2025 20:20
Operator : JC\MD
Sample : PB166989ZHE#13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166989ZHE#13

Quant Time: Mar 07 00:45:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

