

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030625\
 Data File : VN085912.D
 Acq On : 06 Mar 2025 20:44
 Operator : JC\MD
 Sample : PB166989ZHE#14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166989ZHE#14

Quant Time: Mar 07 00:46:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	140497	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	261827	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	252671	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	95320	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	107299	58.907	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 117.820%			
35) Dibromofluoromethane	8.165	113	90609	52.887	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.780%			
50) Toluene-d8	10.565	98	331225	53.137	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 106.280%			
62) 4-Bromofluorobenzene	12.847	95	102833	50.070	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.140%			
Target Compounds						
16) Acetone	4.436	43	32846	56.989	ug/l #	86
20) Methylene Chloride	5.265	84	3602	1.943	ug/l #	73
43) Isopropyl Acetate	8.683	43	63795m	17.866	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030625\
Data File : VN085912.D
Acq On : 06 Mar 2025 20:44
Operator : JC\MD
Sample : PB166989ZHE#14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166989ZHE#14

Quant Time: Mar 07 00:46:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

