

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030722\
 Data File : VN071173.D
 Acq On : 07 Mar 2022 13:41
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Mar 08 05:04:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 04:45:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	721978	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1176046	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1045767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	341214	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	452460	45.726	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.460%
35) Dibromofluoromethane	8.022	113	343931	46.344	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.680%
50) Toluene-d8	10.439	98	1427415	47.771	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.540%
62) 4-Bromofluorobenzene	12.733	95	416028	41.117	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	82.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030722\
Data File : VN071173.D
Acq On : 07 Mar 2022 13:41
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Mar 08 05:04:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 04:45:05 2022
Response via : Initial Calibration

