

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030723\
 Data File : VN076864.D
 Acq On : 07 Mar 2023 11:07
 Operator : JC\MD
 Sample : VN0307MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0307MBL01

Quant Time: Mar 07 22:39:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 07 08:45:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	569153	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	987695	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	853765	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	330120	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	382465	47.930	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.860%
35) Dibromofluoromethane	8.172	113	300251	48.594	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.180%
50) Toluene-d8	10.572	98	1181755	49.643	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.280%
62) 4-Bromofluorobenzene	12.854	95	387496	46.976	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030723\
Data File : VN076864.D
Acq On : 07 Mar 2023 11:07
Operator : JC\MD
Sample : VN0307MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0307MBL01

Quant Time: Mar 07 22:39:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 07 08:45:33 2023
Response via : Initial Calibration

