

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030723\
 Data File : VN076865.D
 Acq On : 07 Mar 2023 11:31
 Operator : JC\MD
 Sample : VN0307WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0307WBL01

Quant Time: Mar 07 22:39:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 07 08:45:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	658605	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	1158127	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	991196	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	388153	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.584	65	450504	48.788	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.580%
35) Dibromofluoromethane	8.172	113	351538	48.521	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.040%
50) Toluene-d8	10.572	98	1380108	49.444	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.880%
62) 4-Bromofluorobenzene	12.854	95	447552	46.272	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030723\
Data File : VN076865.D
Acq On : 07 Mar 2023 11:31
Operator : JC\MD
Sample : VN0307WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0307WBL01

Quant Time: Mar 07 22:39:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 07 08:45:33 2023
Response via : Initial Calibration

