

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030822\
 Data File : VN071184.D
 Acq On : 08 Mar 2022 12:17
 Operator : JC\MD
 Sample : VN0308MBL01
 Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0308MBL01

Quant Time: Mar 09 04:00:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	772388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1236604	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1077192	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	344058	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	474942	44.866	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	89.740%
35) Dibromofluoromethane	8.016	113	367966	47.155	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.300%
50) Toluene-d8	10.439	98	1501206	47.780	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.560%
62) 4-Bromofluorobenzene	12.727	95	430580	40.471	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	80.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030822\
Data File : VN071184.D
Acq On : 08 Mar 2022 12:17
Operator : JC\MD
Sample : VN0308MBL01
Misc : 5.00g/10.0mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0308MBL01

Quant Time: Mar 09 04:00:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

