

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030822\
 Data File : VN071192.D
 Acq On : 08 Mar 2022 16:05
 Operator : JC\MD
 Sample : N1802-01
 Misc : 5.16g/5.0mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 342

Quant Time: Mar 09 04:03:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.075	168	788781	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.963	114	1289503	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1105654	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	431768	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.428	65	494160	45.711	ug/l	-0.01
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.420%
35) Dibromofluoromethane	8.016	113	362204	44.512	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.020%
50) Toluene-d8	10.433	98	1565411	47.780	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.560%
62) 4-Bromofluorobenzene	12.727	95	499001	44.978	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.960%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	11107	2.112	ug/l #	89
18) Methyl Acetate	4.863	43	11889	1.304	ug/l #	69
84) 1,2,4-Trimethylbenzene	13.363	105	18683	0.602	ug/l	99
95) Naphthalene	15.504	128	33063	6.569	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030822\
Data File : VN071192.D
Acq On : 08 Mar 2022 16:05
Operator : JC\MD
Sample : N1802-01
Misc : 5.16g/5.0mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
342

Quant Time: Mar 09 04:03:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

