

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030822\
 Data File : VN071193.D
 Acq On : 08 Mar 2022 16:29
 Operator : JC\MD
 Sample : N1827-13 10X
 Misc : 1.05g/10.0mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RO-291

Quant Time: Mar 09 04:03:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	821216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1319164	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1240191	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	406085	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	507553	45.096	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	90.200%		
35) Dibromofluoromethane	8.016	113	376187	45.191	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	90.380%		
50) Toluene-d8	10.440	98	1639255	48.909	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.820%		
62) 4-Bromofluorobenzene	12.728	95	516961	45.549	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.100%		
Target Compounds						
25) 2-Butanone	7.340	43	9683	1.098	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030822\
Data File : VN071193.D
Acq On : 08 Mar 2022 16:29
Operator : JC\MD
Sample : N1827-13 10X
Misc : 1.05g/10.0mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RO-291

Quant Time: Mar 09 04:03:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

