

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030822\
 Data File : VN071195.D
 Acq On : 08 Mar 2022 17:17
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Mar 09 04:03:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	862517	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1401652	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1230380	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	442112	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	544462	46.058	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.120%
35) Dibromofluoromethane	8.022	113	408647	46.201	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.400%
50) Toluene-d8	10.439	98	1690398	47.467	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.940%
62) 4-Bromofluorobenzene	12.727	95	591326	49.035	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.080%
Target Compounds						
95) Naphthalene	15.503	128	35545	6.644	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030822\
Data File : VN071195.D
Acq On : 08 Mar 2022 17:17
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Mar 09 04:03:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

