

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030822\
 Data File : VN071198.D
 Acq On : 08 Mar 2022 18:30
 Operator : JC\MD
 Sample : PB143183TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143183TB

Quant Time: Mar 09 04:04:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	832805	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1365382	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1238758	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	441013	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	520078	45.565	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	91.140%		
35) Dibromofluoromethane	8.022	113	392069	45.505	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	91.000%		
50) Toluene-d8	10.439	98	1688142	48.663	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	97.320%		
62) 4-Bromofluorobenzene	12.727	95	591009	50.311	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	100.620%		
Target Compounds					Qvalue	
16) Acetone	4.293	43	47984	8.640	ug/l	100
18) Methyl Acetate	4.863	43	8001	1.001	ug/l #	86
20) Methylene Chloride	5.093	84	50685	5.337	ug/l	95
43) Isopropyl Acetate	8.557	43	81567	2.871	ug/l #	90
52) Toluene	10.498	92	45450	1.775	ug/l	90
59) 2-Hexanone	11.086	43	1124203	93.446	ug/l #	49
68) m/p-Xylenes	11.945	106	77223	4.488	ug/l	99
95) Naphthalene	15.504	128	23776	6.124	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030822\
Data File : VN071198.D
Acq On : 08 Mar 2022 18:30
Operator : JC\MD
Sample : PB143183TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143183TB

Quant Time: Mar 09 04:04:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

