

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030822\
 Data File : VN071199.D
 Acq On : 08 Mar 2022 18:54
 Operator : JC\MD
 Sample : N1836-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRE-1282

Quant Time: Mar 09 04:04:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 08 06:52:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	848978	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1369852	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1250514	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	504644	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	521842	44.849	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	89.700%		
35) Dibromofluoromethane	8.022	113	395632	45.768	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	91.540%		
50) Toluene-d8	10.439	98	1676184	48.160	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.320%		
62) 4-Bromofluorobenzene	12.727	95	620432	52.643	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	105.280%		
Target Compounds					Qvalue	
16) Acetone	4.287	43	145868	25.765	ug/l	98
25) 2-Butanone	7.334	43	66910	7.337	ug/l #	89
29) Tetrahydrofuran	7.686	42	39306	6.221	ug/l	91
68) m/p-Xylenes	11.945	106	59753	3.440	ug/l	99
69) o-Xylene	12.280	106	44252	2.541	ug/l	95
78) n-propylbenzene	12.916	91	74939	1.436	ug/l	95
80) 1,3,5-Trimethylbenzene	13.057	105	164227	4.339	ug/l	99
84) 1,2,4-Trimethylbenzene	13.363	105	892931	24.631	ug/l	99
89) n-Butylbenzene	13.939	91	65385	2.368	ug/l #	73
95) Naphthalene	15.498	128	256676	15.038	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030822\
Data File : VN071199.D
Acq On : 08 Mar 2022 18:54
Operator : JC\MD
Sample : N1836-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRE-1282

Quant Time: Mar 09 04:04:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030722W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 08 06:52:46 2022
Response via : Initial Calibration

