

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030823\
 Data File : VN076879.D
 Acq On : 08 Mar 2023 12:20
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Mar 09 03:21:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 07 08:45:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	529180	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	903237	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.866	117	776548	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	286226	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	353535	47.651	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.300%
35) Dibromofluoromethane	8.172	113	277716	49.149	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.300%
50) Toluene-d8	10.572	98	1080135	49.617	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.240%
62) 4-Bromofluorobenzene	12.854	95	338245	44.840	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030823\
Data File : VN076879.D
Acq On : 08 Mar 2023 12:20
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Mar 09 03:21:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 07 08:45:33 2023
Response via : Initial Calibration

