

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030823\
 Data File : VN076888.D
 Acq On : 08 Mar 2023 19:59
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Mar 09 03:48:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N030823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 09 03:46:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.825	128	44655	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.107	114	314811	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.866	117	342405	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.584	65	126762	30.792	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.633%
60) 4-Bromofluorobenzene	12.854	95	173772	28.423	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.733%
63) Toluene-d8	10.572	98	346908	28.913	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.367%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030823\
Data File : VN076888.D
Acq On : 08 Mar 2023 19:59
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Mar 09 03:48:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N030823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 09 03:46:27 2023
Response via : Initial Calibration

