

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030823\
 Data File : VN076874.D
 Acq On : 08 Mar 2023 09:59
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 09 03:20:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 07 08:45:33 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.582	0.571	1.9	99	0.00
3 P	Chloromethane	0.765	0.646	15.6	89	0.00
4 C	Vinyl Chloride	0.701	0.601	14.3#	87	0.00
5 T	Bromomethane	0.461	0.374	18.9	88	0.00
6 T	Chloroethane	0.477	0.367	23.1	88	0.00
7 T	Trichlorofluoromethane	1.023	0.994	2.8	98	0.00
8 T	Diethyl Ether	0.394	0.397	-0.8	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.598	0.596	0.3	103	0.00
10 T	Methyl Iodide	0.624	0.711	-13.9	98	0.00
11 T	Tert butyl alcohol	0.104	0.092	11.5	89	0.00
12 CM	1,1-Dichloroethene	0.558	0.570	-2.2#	101	0.00
13 T	Acrolein	0.135	0.126	6.7	93	0.00
14 T	Allyl chloride	0.932	1.012	-8.6	106	0.00
15 T	Acrylonitrile	0.335	0.322	3.9	94	0.00
16 T	Acetone	0.306	0.307	-0.3	107	0.00
17 T	Carbon Disulfide	1.515	1.513	0.1	99	0.00
18 T	Methyl Acetate	1.049	0.983	6.3	94	0.00
19 T	Methyl tert-butyl Ether	1.913	1.926	-0.7	98	0.00
20 T	Methylene Chloride	0.714	0.660	7.6	98	0.00
21 T	trans-1,2-Dichloroethene	0.603	0.611	-1.3	100	0.00
22 T	Diisopropyl ether	2.096	2.105	-0.4	96	0.00
23 T	Vinyl Acetate	1.335	1.351	-1.2	94	0.00
24 P	1,1-Dichloroethane	1.195	1.191	0.3	100	0.00
25 T	2-Butanone	0.447	0.433	3.1	95	0.00
26 T	2,2-Dichloropropane	0.875	0.904	-3.3	100	0.00
27 T	cis-1,2-Dichloroethene	0.707	0.704	0.4	98	0.00
28 T	Bromochloromethane	0.505	0.485	4.0	99	0.00
29 T	Tetrahydrofuran	0.288	0.275	4.5	91	0.00
30 C	Chloroform	1.187	1.182	0.4#	100	0.00
31 T	Cyclohexane	1.184	1.087	8.2	96	0.00
32 T	1,1,1-Trichloroethane	1.028	1.022	0.6	98	0.00
33 S	1,2-Dichloroethane-d4	0.701	0.609	13.1	89	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.313	0.290	7.3	92	0.00
36 T	1,1-Dichloropropene	0.503	0.526	-4.6	99	0.00
37 T	Ethyl Acetate	0.516	0.491	4.8	91	0.00
38 T	Carbon Tetrachloride	0.511	0.527	-3.1	100	0.00
39 T	Methylcyclohexane	0.620	0.662	-6.8	98	0.00
40 TM	Benzene	1.531	1.594	-4.1	100	0.00
41 T	Methacrylonitrile	0.282	0.288	-2.1	94	0.00
42 TM	1,2-Dichloroethane	0.533	0.521	2.3	94	0.00
43 T	Isopropyl Acetate	0.790	0.792	-0.3	92	0.00
44 TM	Trichloroethene	0.363	0.388	-6.9	102	0.00
45 C	1,2-Dichloropropane	0.396	0.407	-2.8#	97	0.00
46 T	Dibromomethane	0.251	0.258	-2.8	100	0.00
47 T	Bromodichloromethane	0.517	0.544	-5.2	98	0.00
48 T	Methyl methacrylate	0.384	0.387	-0.8	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030823\
 Data File : VN076874.D
 Acq On : 08 Mar 2023 09:59
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 09 03:20:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 07 08:45:33 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	91	0.00
50 S	Toluene-d8	1.205	1.160	3.7	92	0.00
51 T	4-Methyl-2-Pentanone	0.498	0.500	-0.4	90	0.00
52 CM	Toluene	0.912	0.991	-8.7#	100	0.00
53 T	t-1,3-Dichloropropene	0.501	0.561	-12.0	98	0.00
54 T	cis-1,3-Dichloropropene	0.575	0.632	-9.9	99	0.00
55 T	1,1,2-Trichloroethane	0.363	0.376	-3.6	99	0.00
56 T	Ethyl methacrylate	0.540	0.590	-9.3	95	0.00
57 T	1,3-Dichloropropane	0.630	0.661	-4.9	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.187	0.188	-0.5	88	0.00
59 T	2-Hexanone	0.361	0.365	-1.1	92	0.00
60 T	Dibromochloromethane	0.357	0.393	-10.1	99	0.00
61 T	1,2-Dibromoethane	0.349	0.370	-6.0	99	0.00
62 S	4-Bromofluorobenzene	0.418	0.412	1.4	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.372	0.430	-15.6	109	0.00
65 PM	Chlorobenzene	1.081	1.100	-1.8	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.391	0.407	-4.1	100	0.00
67 C	Ethyl Benzene	1.962	2.107	-7.4#	99	0.00
68 T	m/p-Xylenes	0.738	0.800	-8.4	100	0.00
69 T	o-Xylene	0.727	0.785	-8.0	99	0.00
70 T	Styrene	1.140	1.288	-13.0	99	0.00
71 P	Bromoform	0.266	0.294	-10.5	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	4.358	4.428	-1.6	99	0.00
74 T	N-amyl acetate	1.705	1.640	3.8	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.397	1.222	12.5	94	0.00
76 T	1,2,3-Trichloropropane	1.261	1.152	8.6	96	0.00
77 T	Bromobenzene	0.964	0.972	-0.8	102	0.00
78 T	n-propylbenzene	4.970	5.285	-6.3	100	0.00
79 T	2-Chlorotoluene	3.188	3.169	0.6	100	0.00
80 T	1,3,5-Trimethylbenzene	3.653	3.809	-4.3	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.382	0.390	-2.1	98	0.00
82 T	4-Chlorotoluene	3.023	3.086	-2.1	100	0.00
83 T	tert-Butylbenzene	3.121	3.309	-6.0	103	0.00
84 T	1,2,4-Trimethylbenzene	3.603	3.802	-5.5	100	0.00
85 T	sec-Butylbenzene	4.358	4.723	-8.4	101	0.00
86 T	p-Isopropyltoluene	3.486	3.851	-10.5	102	0.00
87 T	1,3-Dichlorobenzene	1.761	1.831	-4.0	102	0.00
88 T	1,4-Dichlorobenzene	1.788	1.813	-1.4	103	0.00
89 T	n-Butylbenzene	3.011	3.355	-11.4	102	0.00
90 T	Hexachloroethane	0.630	0.650	-3.2	99	0.00
91 T	1,2-Dichlorobenzene	1.784	1.803	-1.1	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.265	0.247	6.8	92	0.00
93 T	1,2,4-Trichlorobenzene	0.892	1.004	-12.6	102	0.00
94 T	Hexachlorobutadiene	0.465	0.501	-7.7	108	0.00
95 T	Naphthalene	2.859	3.149	-10.1	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030823\
Data File : VN076874.D
Acq On : 08 Mar 2023 09:59
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 09 03:20:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030623W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 07 08:45:33 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.877	0.986	-12.4	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6