

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030824\
 Data File : VN081337.D
 Acq On : 08 Mar 2024 10:29
 Operator : JC\MD
 Sample : VN0308MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0308MBL01

Quant Time: Mar 09 00:53:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	315896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	566676	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	490759	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	198525	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	223878	49.015	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.040%
35) Dibromofluoromethane	8.171	113	176933	50.962	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.920%
50) Toluene-d8	10.570	98	671047	51.704	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.400%
62) 4-Bromofluorobenzene	12.847	95	212943	46.422	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	92.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030824\
Data File : VN081337.D
Acq On : 08 Mar 2024 10:29
Operator : JC\MD
Sample : VN0308MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0308MBL01

Quant Time: Mar 09 00:53:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

