

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030824\
 Data File : VN081342.D
 Acq On : 08 Mar 2024 13:26
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Mar 09 00:56:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	302505	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	557004	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	481525	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	189845	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	223290	51.051	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.100%
35) Dibromofluoromethane	8.165	113	173228	50.762	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.520%
50) Toluene-d8	10.571	98	656005	51.423	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.840%
62) 4-Bromofluorobenzene	12.853	95	204528	45.362	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	90.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030824\
Data File : VN081342.D
Acq On : 08 Mar 2024 13:26
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Mar 09 00:56:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

