

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030824\
 Data File : VN081345.D
 Acq On : 08 Mar 2024 14:38
 Operator : JC\MD
 Sample : P1717-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Mar 09 00:57:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 06 03:12:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	298707	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	542873	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	473841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	182961	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	222127	51.431	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.860%
35) Dibromofluoromethane	8.171	113	172780	51.948	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.900%
50) Toluene-d8	10.571	98	643620	51.765	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.540%
62) 4-Bromofluorobenzene	12.853	95	206293	46.944	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	93.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN030824\
Data File : VN081345.D
Acq On : 08 Mar 2024 14:38
Operator : JC\MD
Sample : P1717-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Mar 09 00:57:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N030524W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 06 03:12:57 2024
Response via : Initial Calibration

